

LB114731


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50114



A= 0.0000e+000

B= 1.9802e-004

C= -4.8748e-002

Rho= 0.9997584

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o
0.0	0.000	-0.024	-0.024	125	0.000	125					-
0.2	0.200	0.165	-0.035	1079	0.0 %	1079					-18
2.5	2.500	2.477	-0.023	12757	0.0 %	12757					-1
5.0	5.000	5.110	0.110	26053	0.0 %	26053					2
7.5	7.500	7.597	0.097	38609	0.0 %	38609					1
10.0	10.000	9.875	-0.125	50114	0.0 %	50114					-1

LB114731 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Jun 2021 09:36:38

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	125	-	0.00007471B	PPB	02 Jun 2021 09:57:01	S	Std
0.2 - 1	50.2	1079	-	0.20007471B	PPB	02 Jun 2021 09:59:17	S	Std
2.5 - 1	52.5	12757	-	2.50007471B	PPB	02 Jun 2021 10:01:34	S	Std
5.0 - 1	55	26053	-	5.00007471B	PPB	02 Jun 2021 10:03:51	S	Std
7.5 - 1	57.5	38609	-	7.50007471B	PPB	02 Jun 2021 10:06:07	S	Std
10.0 - 1	59.5	50114	-	10.00007471B	PPB	02 Jun 2021 10:08:24	S	Std
ICV44 - 1	ICV44	20133	3.9380	-7471B	PPB	02 Jun 2021 10:12:06	U	SMPL
ICB44 - 1	ICB44	-53	-0.0592	-7471B	PPB	02 Jun 2021 10:14:20	U	SMPL
CCV01 - 1	CCV01	25694	5.0392	-7471B	PPB	02 Jun 2021 10:16:36	U	SMPL
CCB01 - 1	CCB01	-75	-0.0636	-7471B	PPB	02 Jun 2021 10:18:51	U	SMPL
CRA - 1	CRA	1057	0.1606	-7471B	PPB	02 Jun 2021 10:21:06	U	SMPL
HighStd - 1	HighStd	50486	9.9485	-7471B	PPB	02 Jun 2021 10:25:27	U	SMPL
ChkStd - 1	ChkStd	35491	6.9792	-7471B	PPB	02 Jun 2021 10:27:41	U	SMPL
PB136775BL - 1	PBS	-169	-0.0822	-7471B	PPB	02 Jun 2021 10:29:58	U	SMPL
PB136775BS - 1	LCSS	18466	3.6079	-7471B	PPB	02 Jun 2021 10:32:15	U	SMPL
M2546-01 - 1	COMPOSITE-SAMPLE-LOAMY-TOPSOIL-PILE-2	1238	0.1964	-7471B	PPB	02 Jun 2021 10:34:31	U	SMPL
M2547-01 - 1	COMPOSITE- SAMPLE-LOAMY-TOPSOIL-PILE-1	1731	0.2940	-7471B	PPB	02 Jun 2021 10:36:48	U	SMPL
M2560-01 - 1	S1-COMP-A-B-C-D	942	0.1378	-7471B	PPB	02 Jun 2021 10:39:04	U	SMPL
M2565-01 - 1	CL-02-052821	960	0.1414	-7471B	PPB	02 Jun 2021 10:41:22	U	SMPL
CCV02 - 1	CCV02	25598	5.0202	-7471B	PPB	02 Jun 2021 10:43:36	U	SMPL
CCB02 - 1	CCB02	-73	-0.0632	-7471B	PPB	02 Jun 2021 10:45:52	U	SMPL
M2571-01 - 1	SA-1	81	-0.0327	-7471B	PPB	02 Jun 2021 10:48:07	U	SMPL
M2571-01DUP - 1	SA-1DUP	76	-0.0337	-7471B	PPB	02 Jun 2021 10:50:23	U	SMPL
M2571-01MS - 1	SA-1MS	17089	3.3352	-7471B	PPB	02 Jun 2021 10:52:38	U	SMPL
M2571-01MSD - 1	SA-1MSD	18840	3.6819	-7471B	PPB	02 Jun 2021 10:54:53	U	SMPL
M2571-01LX5 - 1		-59	-0.0604	-7471B	PPB	02 Jun 2021 10:57:10	U	SMPL
M2571-01A - 1		20363	3.9835	-7471B	PPB	02 Jun 2021 10:59:27	U	SMPL
CCV03 - 1	CCV03	25908	5.0816	-7471B	PPB	02 Jun 2021 11:01:43	U	SMPL
CCB03 - 1	CCB03	-85	-0.0656	-7471B	PPB	02 Jun 2021 11:03:59	U	SMPL