



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114740

Slope : 97.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/02/2021	12:10
2	CAL2	1	Water	NA	NA	20.3	7.00	06/02/2021	12:11
3	CAL3	1	Water	NA	NA	20.3	10.03	06/02/2021	12:12
4	ICV	1	Water	NA	NA	20.4	6.99	06/02/2021	12:15
5	CCV1	1	Water	NA	NA	20.4	2.01	06/02/2021	12:16
6	M2580-18	1	Solid	20.02	20	23.4	9.10	06/02/2021	12:18
7	M2580-18DUP	1	Solid	20.03	20	23.5	9.11	06/02/2021	12:19
8	M2580-19	1	Solid	20.04	20	23.0	10.77	06/02/2021	12:20
9	M2580-20	1	Solid	20.01	20	23.0	9.54	06/02/2021	12:22
10	M2580-21	1	Solid	20.03	20	23.1	10.07	06/02/2021	12:25
11	M2580-22	1	Solid	20.05	20	23.2	10.70	06/02/2021	12:26
12	M2580-23	1	Solid	20.03	20	23.1	11.19	06/02/2021	12:29
13	M2580-24	1	Solid	20.04	20	23.1	11.34	06/02/2021	12:30
14	M2580-25	1	Solid	20.02	20	23.0	10.55	06/02/2021	12:32
15	M2580-26	1	Solid	20.03	20	23.0	11.38	06/02/2021	12:35
16	CCV2	1	Water	NA	NA	20.2	12.02	06/02/2021	12:36
17	M2580-27	1	Solid	20.03	20	23.2	9.33	06/02/2021	12:39
18	M2580-28	1	Solid	20.04	20	22.9	10.62	06/02/2021	12:40
19	M2580-29	1	Solid	20.02	20	22.9	11.85	06/02/2021	12:42
20	M2580-30	1	Solid	20.03	20	23.0	10.08	06/02/2021	12:45

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M2580-31	1	Solid	20.02	20	23.0	8.86	06/02/2021	12:47
22	M2580-32	1	Solid	20.02	20	23.0	10.61	06/02/2021	12:49
23	M2580-33	1	Solid	20.03	20	22.8	9.50	06/02/2021	12:50
24	M2580-34	1	Solid	20.04	20	22.9	10.90	06/02/2021	12:51
25	CCV3	1	Water	NA	NA	20.5	2.01	06/02/2021	13:00

18112740

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M2580 WorkList ID : 149430 Department : Wet-Chemistry Date : 06-02-2021 11:52:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/01/2021	Solid	M2580-18	Corrosivity	Cool 4 deg C	WALS01	D62	18-B1(0-2)	06/01/2021	9045D
06/01/2021	Solid	M2580-19	Corrosivity	Cool 4 deg C	WALS01	D62	18-B1(2-4)	06/01/2021	9045D
06/01/2021	Solid	M2580-20	Corrosivity	Cool 4 deg C	WALS01	D62	18-B1(4-6)	06/01/2021	9045D
06/01/2021	Solid	M2580-21	Corrosivity	Cool 4 deg C	WALS01	D62	18-B2(0-2)	06/01/2021	9045D
06/01/2021	Solid	M2580-22	Corrosivity	Cool 4 deg C	WALS01	D62	18-B2(2-4)	06/01/2021	9045D
06/01/2021	Solid	M2580-23	Corrosivity	Cool 4 deg C	WALS01	D62	18-B2(4-6)	06/01/2021	9045D
06/01/2021	Solid	M2580-24	Corrosivity	Cool 4 deg C	WALS01	D62	18-B3(0-2)	06/01/2021	9045D
06/01/2021	Solid	M2580-25	Corrosivity	Cool 4 deg C	WALS01	D62	18-B3(2-4)	06/01/2021	9045D
06/01/2021	Solid	M2580-26	Corrosivity	Cool 4 deg C	WALS01	D62	18-B3(4-6)	06/01/2021	9045D
06/01/2021	Solid	M2580-27	Corrosivity	Cool 4 deg C	WALS01	D62	18-B4(0-2)	06/01/2021	9045D
06/01/2021	Solid	M2580-28	Corrosivity	Cool 4 deg C	WALS01	D62	18-B4(2-4)	06/01/2021	9045D
06/01/2021	Solid	M2580-29	Corrosivity	Cool 4 deg C	WALS01	D62	18-B4(4-6)	06/01/2021	9045D
06/01/2021	Solid	M2580-30	Corrosivity	Cool 4 deg C	WALS01	D62	18-B5(0-2)	06/01/2021	9045D
06/01/2021	Solid	M2580-31	Corrosivity	Cool 4 deg C	WALS01	D62	18-B5(2-4)	06/01/2021	9045D
06/01/2021	Solid	M2580-32	Corrosivity	Cool 4 deg C	WALS01	D62	18-B5(4-6)	06/01/2021	9045D
06/01/2021	Solid	M2580-33	Corrosivity	Cool 4 deg C	WALS01	D62	18-B6(0-6)	06/01/2021	9045D
06/01/2021	Solid	M2580-34	Corrosivity	Cool 4 deg C	WALS01	D62	18-B6(6-12)	06/01/2021	9045D

Date/Time 06/02/21 12:00
 Raw Sample Received by: Jo (wet Chem)
 Raw Sample Relinquished by: KS (wet Chem)

Date/Time 06/02/21 16:30
 Raw Sample Received by: KS (wet Chem)
 Raw Sample Relinquished by: Jo (wet Chem)