



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114743

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.5	4.01	06/02/2021	15:32
2	CAL2	1	Water	NA	NA	22.8	7.01	06/02/2021	15:34
3	CAL3	1	Water	NA	NA	22.7	10.02	06/02/2021	15:36
4	ICV	1	Water	NA	NA	22.8	7.02	06/02/2021	15:39
5	CCV1	1	Water	NA	NA	22.8	2.02	06/02/2021	15:41
6	M2590-01	1	Solid	20.08	20	22.9	11.84	06/02/2021	15:44
7	M2590-01DUP	1	Solid	20.08	20	22.8	11.86	06/02/2021	15:47
8	CCV2	1	Water	NA	NA	22.8	12.02	06/02/2021	15:50

LB114743

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH-60221 WorkList ID : 149434 Department : Wet-Chemistry Date : 06-02-2021 15:21:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/02/2021	Solid	M2590-01	pH	Cool 4 deg C	PARS15	F11	JTY-FTS	06/02/2021	9045D

Date/Time 06/02/2021 15:25
Raw Sample Received by: AM (Wet-Chem)
Raw Sample Relinquished by: CP SM

Date/Time 06/02/2021 15:55
Raw Sample Received by: CP SM
Raw Sample Relinquished by: AM (Wet-Chem)