



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB114746

Analyst By : Alexander

Supervisor Review By : apatel

Slope : 98.7

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	24.7	4.01	06/02/2021	16:35
2	CAL2	1	Water	NA	NA	24.6	7.01	06/02/2021	16:37
3	CAL3	1	Water	NA	NA	24.8	10.03	06/02/2021	16:38
4	ICV	1	Water	NA	NA	24.8	6.99	06/02/2021	16:40
5	CCV1	1	Water	NA	NA	24.9	2.01	06/02/2021	16:41
6	M2583-06	1	Water	NA	NA	25.0	6.58	06/02/2021	16:43
7	M2583-06DUP	1	Water	NA	NA	25.0	6.57	06/02/2021	16:45
8	M2583-07	1	Water	NA	NA	25.0	6.59	06/02/2021	16:47
9	M2583-08	1	Water	NA	NA	25.0	6.65	06/02/2021	16:49
10	M2583-09	1	Water	NA	NA	24.9	6.61	06/02/2021	16:52
11	M2583-10	1	Water	NA	NA	24.9	7.08	06/02/2021	16:54
12	M2583-11	1	Water	NA	NA	24.7	7.53	06/02/2021	16:56
13	M2583-12	1	Water	NA	NA	24.6	7.24	06/02/2021	16:58
14	M2583-13	1	Water	NA	NA	24.4	6.62	06/02/2021	16:59
15	M2583-14	1	Water	NA	NA	24.5	6.83	06/02/2021	17:01
16	CCV2	1	Water	NA	NA	24.9	12.03	06/02/2021	17:02
17	M2583-15	1	Water	NA	NA	25.0	6.57	06/02/2021	17:04
18	M2583-16	1	Water	NA	NA	24.8	6.63	06/02/2021	17:06
19	M2583-16DUP	1	Water	NA	NA	24.8	6.66	06/02/2021	17:08
20	M2583-17	1	Water	NA	NA	24.9	6.73	06/02/2021	17:11

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M2583-18	1	Water	NA	NA	24.8	6.78	06/02/2021	17:14
22	CCV3	1	Water	NA	NA	24.8	2.02	06/02/2021	17:16

LB114746

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH-60221-2 WorkList ID : 149435 Department : Wet-Chemistry Date : 06-02-2021 16:16:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/02/2021	Water	M2583-06	pH	Cool 4 deg C	PSEG03	G11	41833	06/02/2021	9040C
06/02/2021	Water	M2583-07	pH	Cool 4 deg C	PSEG03	G11	41834	06/02/2021	9040C
06/02/2021	Water	M2583-08	pH	Cool 4 deg C	PSEG03	G11	41835	06/02/2021	9040C
06/02/2021	Water	M2583-09	pH	Cool 4 deg C	PSEG03	G11	41836	06/02/2021	9040C
06/02/2021	Water	M2583-10	pH	Cool 4 deg C	PSEG03	G11	41837	06/02/2021	9040C
06/02/2021	Water	M2583-11	pH	Cool 4 deg C	PSEG03	G11	41838	06/02/2021	9040C
06/02/2021	Water	M2583-12	pH	Cool 4 deg C	PSEG03	G11	41839	06/02/2021	9040C
06/02/2021	Water	M2583-13	pH	Cool 4 deg C	PSEG03	G11	41840	06/02/2021	9040C
06/02/2021	Water	M2583-14	pH	Cool 4 deg C	PSEG03	G11	41841	06/02/2021	9040C
06/02/2021	Water	M2583-15	pH	Cool 4 deg C	PSEG03	G11	41854	06/02/2021	9040C
06/02/2021	Water	M2583-16	pH	Cool 4 deg C	PSEG03	G11	41855	06/02/2021	9040C
06/02/2021	Water	M2583-17	pH	Cool 4 deg C	PSEG03	G11	41856	06/02/2021	9040C
06/02/2021	Water	M2583-18	pH	Cool 4 deg C	PSEG03	G11	41857	06/02/2021	9040C

Date/Time 06/02/2021 16:20
Raw Sample Received by: AM (wet-chem)
Raw Sample Relinquished by: KS (wet-chem)

Date/Time 06/02/2021 17:30
Raw Sample Received by: KS (wet-chem)
Raw Sample Relinquished by: AM (wet-chem)