

LB114750


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53152



A= 0.0000e+000

B= 1.8757e-004

C= 2.1969e-002

Rho= 0.9999256

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.031	0.031	50	0.000	50					15
0.2	0.200	0.229	0.029	1103	0.0 %	1103					3
2.5	2.500	2.431	-0.069	12842	0.0 %	12842					1
5.0	5.000	4.957	-0.043	26309	0.0 %	26309					0
7.5	7.500	7.560	0.060	40188	0.0 %	40188					
10.0	10.000	9.992	-0.008	53152	0.0 %	53152					

LB114750 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 03 Jun 2021 09:18:31				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	50	-	0.0000	7471B	PPB	03 Jun 2021 09:25:59	S Std
0.2 - 1	502	1103	-	0.2000	7471B	PPB	03 Jun 2021 09:28:15	S Std
2.5 - 1	52.5	12842	-	2.5000	7471B	PPB	03 Jun 2021 09:30:32	S Std
5.0 - 1	55	26309	-	5.0000	7471B	PPB	03 Jun 2021 09:32:49	S Std
7.5 - 1	57.5	40188	-	7.5000	7471B	PPB	03 Jun 2021 09:35:05	S Std
10.0 - 1	510	53152	-	10.0000	7471B	PPB	03 Jun 2021 09:37:22	S Std
ICV46 - 1	ICV46	20522	3.8714	-	7471B	PPB	03 Jun 2021 09:40:08	U SMPL
ICB46 - 1	ICB46	-105	0.0023	-	7471B	PPB	03 Jun 2021 09:42:23	U SMPL
CCV08 - 1	CCV08	27325	5.1474	-	7471B	PPB	03 Jun 2021 09:44:38	U SMPL
CCB08 - 1	CCB08	-111	0.0011	-	7471B	PPB	03 Jun 2021 09:46:53	U SMPL
CRA - 1	CRA	1030	0.2152	-	7471B	PPB	03 Jun 2021 09:49:09	U SMPL
HighStd - 1	HighStd	53088	9.9799	-	7471B	PPB	03 Jun 2021 09:51:23	U SMPL
ChkStd - 1	ChkStd	36501	6.8686	-	7471B	PPB	03 Jun 2021 09:53:38	U SMPL
PB136788BL - 1	PBS	-168	-0.0095	-	7471B	PPB	03 Jun 2021 09:55:55	U SMPL
PB136788BS - 1	LCSS	20413	3.8509	-	7471B	PPB	03 Jun 2021 09:58:12	U SMPL
M2580-01 - 1	18-B1(0-2)	15863	2.9975	-	7471B	PPB	03 Jun 2021 10:05:24	U SMPL
M2580-01DUP - 1	18-B1(0-2)DUP	15569	2.9423	-	7471B	PPB	03 Jun 2021 10:07:40	U SMPL
M2580-01MS - 1	18-B1(0-2)MS	36015	6.7775	-	7471B	PPB	03 Jun 2021 10:09:56	U SMPL
M2580-01MSD - 1	18-B1(0-2)MSD	33224	6.2539	-	7471B	PPB	03 Jun 2021 10:12:13	U SMPL
M2580-02 - 1	18-B1(2-4)	10871	2.0611	-	7471B	PPB	03 Jun 2021 10:14:29	U SMPL
CCV09 - 1	CCV09	27631	5.2048	-	7471B	PPB	03 Jun 2021 10:16:45	U SMPL
CCB09 - 1	CCB09	-93	0.0045	-	7471B	PPB	03 Jun 2021 10:19:01	U SMPL
M2580-03 - 1	18-B1(4-6)	11646	2.2065	-	7471B	PPB	03 Jun 2021 10:21:16	U SMPL
M2580-04 - 1	18-B2(0-2)	17575	3.3186	-	7471B	PPB	03 Jun 2021 10:23:32	U SMPL
M2580-05 - 1	18-B2(2-4)	15360	2.9031	-	7471B	PPB	03 Jun 2021 10:25:46	U SMPL
M2580-06 - 1	18-B2(4-6)	11525	2.1838	-	7471B	PPB	03 Jun 2021 10:28:02	U SMPL
M2580-07 - 1	18-B3(0-2)	3716	0.7190	-	7471B	PPB	03 Jun 2021 10:30:17	U SMPL
M2580-08 - 1	18-B3(2-4)	7795	1.4841	-	7471B	PPB	03 Jun 2021 10:32:34	U SMPL
M2580-09 - 1	18-B3(4-6)	11196	2.1221	-	7471B	PPB	03 Jun 2021 10:34:50	U SMPL
M2580-10 - 1	18-B4(0-2)	4653	0.8948	-	7471B	PPB	03 Jun 2021 10:37:07	U SMPL
M2580-11 - 1	18-B4(2-4)	31085	5.8527	-	7471B	PPB	03 Jun 2021 10:39:24	U SMPL
M2580-12 - 1	18-B4(4-6)	7700	1.4663	-	7471B	PPB	03 Jun 2021 10:41:40	U SMPL
CCV10 - 1	CCV10	27354	5.1529	-	7471B	PPB	03 Jun 2021 10:43:56	U SMPL
CCB10 - 1	CCB10	-88	0.0055	-	7471B	PPB	03 Jun 2021 10:46:11	U SMPL
M2580-13 - 1	18-B5(0-2)	42052	7.9098	-	7471B	PPB	03 Jun 2021 10:48:27	U SMPL
M2580-14 - 1	18-B5(2-4)	420	0.1008	-	7471B	PPB	03 Jun 2021 10:50:42	U SMPL
M2580-15 - 1	18-B5(4-6)	25407	4.7877	-	7471B	PPB	03 Jun 2021 10:52:59	U SMPL
M2580-16 - 1	18-B6(0-6)	26887	5.0653	-	7471B	PPB	03 Jun 2021 10:55:15	U SMPL
M2580-17 - 1	18-B6(6-12)	19117	3.6078	-	7471B	PPB	03 Jun 2021 10:57:31	U SMPL
PB136789BL - 1	PBS	-40	0.0145	-	7471B	PPB	03 Jun 2021 10:59:48	U SMPL
PB136789BS - 1	LCSS	19612	3.7007	-	7471B	PPB	03 Jun 2021 11:02:04	U SMPL
M2573-01 - 1	OR-02-060121	918	0.1942	-	7471B	PPB	03 Jun 2021 11:04:21	U SMPL
M2574-01 - 1	NB-07-060121	397	0.0964	-	7471B	PPB	03 Jun 2021 11:06:39	U SMPL
M2581-01 - 1	TR-05-060221	2846	0.5558	-	7471B	PPB	03 Jun 2021 11:08:54	U SMPL
CCV11 - 1	CCV11	27157	5.1159	-	7471B	PPB	03 Jun 2021 11:11:10	U SMPL
CCB11 - 1	CCB11	-93	0.0045	-	7471B	PPB	03 Jun 2021 11:13:25	U SMPL
M2582-01 - 1	RO-1-VNJ-201	1992	0.3956	-	7471B	PPB	03 Jun 2021 11:15:40	U SMPL
M2583-03 - 1	41122	189	0.0574	-	7471B	PPB	03 Jun 2021 11:17:56	U SMPL
M2583-04 - 1	40955	264	0.0715	-	7471B	PPB	03 Jun 2021 11:20:11	U SMPL
M2583-05 - 1	30785	351	0.0878	-	7471B	PPB	03 Jun 2021 11:22:27	U SMPL
M2584-01 - 1	EO-02-060221	1396	0.2838	-	7471B	PPB	03 Jun 2021 11:24:43	U SMPL
M2584-03 - 1	EO-03-060221	2231	0.4404	-	7471B	PPB	03 Jun 2021 11:26:59	U SMPL
M2585-02 - 1	ETGI-363	3979	0.7683	-	7471B	PPB	03 Jun 2021 11:29:15	U SMPL
M2587-01 - 1	FK-GF-02-073-A	5899	1.1285	-	7471B	PPB	03 Jun 2021 11:31:31	U SMPL
M2587-03 - 1	FK-GF-02-073-B	5894	1.1275	-	7471B	PPB	03 Jun 2021 11:33:48	U SMPL
M2588-01 - 1	FK-GF-02-074-A	6017	1.1506	-	7471B	PPB	03 Jun 2021 11:36:04	U SMPL
CCV12 - 1	CCV12	25685	4.8398	-	7471B	PPB	03 Jun 2021 11:38:19	U SMPL
CCB12 - 1	CCB12	-85	0.0060	-	7471B	PPB	03 Jun 2021 11:40:34	U SMPL
M2588-03 - 1	FK-GF-02-074-B	5341	1.0238	-	7471B	PPB	03 Jun 2021 11:42:50	U SMPL
M2588-03DUP - 1	FK-GF-02-074-BDUP	5658	1.0833	-	7471B	PPB	03 Jun 2021 11:45:06	U SMPL
M2588-03MS - 1	FK-GF-02-074-BMS	25471	4.7997	-	7471B	PPB	03 Jun 2021 11:47:21	U SMPL
M2588-03MSD - 1	FK-GF-02-074-BMSD	25657	4.8346	-	7471B	PPB	03 Jun 2021 11:49:37	U SMPL
M2591-01 - 1	S-1	1867	0.3722	-	7471B	PPB	03 Jun 2021 11:51:54	U SMPL
M2580-01LX5 - 1		3044	0.5929	-	7471B	PPB	03 Jun 2021 11:54:12	U SMPL
M2580-01A - 1		35542	6.6887	-	7471B	PPB	03 Jun 2021 11:56:28	U SMPL
M2588-03LX5 - 1		701	0.1535	-	7471B	PPB	03 Jun 2021 11:58:44	U SMPL
M2588-03A - 1		26632	5.0174	-	7471B	PPB	03 Jun 2021 13:00:03	U SMPL
CCV13 - 1	CCV13	25185	4.7460	-	7471B	PPB	03 Jun 2021 13:02:19	U SMPL
CCB13 - 1	CCB13	-77	0.0075	-	7471B	PPB	03 Jun 2021 13:04:35	U SMPL