



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114754

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/03/2021	11:40
2	CAL2	1	Water	NA	NA	20.3	7.00	06/03/2021	11:41
3	CAL3	1	Water	NA	NA	20.3	10.02	06/03/2021	11:42
4	ICV	1	Water	NA	NA	20.4	6.99	06/03/2021	11:44
5	CCV1	1	Water	NA	NA	20.3	2.01	06/03/2021	11:45
6	M2593-01	1	Water	NA	NA	18.7	6.67	06/03/2021	11:48
7	M2593-01DUP	1	Water	NA	NA	18.8	6.68	06/03/2021	11:49
8	CCV2	1	Water	NA	NA	20.3	12.02	06/03/2021	11:50

VB 114754

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2593 WorkList ID : 149448 Department : Wet-Chemistry Date : 06-03-2021 11:34:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/02/2021	Water	M2593-01	pH ✓	Cool 4 deg C	APT101	F11	WGGM-21-WWW-TSF-023	06/02/2021	SM 4500-H B

Date/Time 06/03/21 11:35
Raw Sample Received by: CS (Wetchem)
Raw Sample Relinquished by: CS (Wetchem)

Date/Time 06/03/21 12:35
Raw Sample Received by: CS (Wetchem)
Raw Sample Relinquished by: CS (Wetchem)