



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114762

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/03/2021	17:10
2	CAL2	1	Water	NA	NA	20.3	7.00	06/03/2021	17:11
3	CAL3	1	Water	NA	NA	20.3	10.02	06/03/2021	17:12
4	ICV	1	Water	NA	NA	20.3	6.99	06/03/2021	17:15
5	CCV1	1	Water	NA	NA	20.3	2.01	06/03/2021	17:16
6	M2601-03	1	Solid	20.04	20	23.9	6.36	06/03/2021	17:26
7	M2601-05	1	Solid	20.03	20	23.9	6.48	06/03/2021	17:28
8	M2601-07	1	Solid	20.04	20	23.9	6.75	06/03/2021	17:29
9	M2601-09	1	Solid	20.03	20	23.8	6.29	06/03/2021	17:30
10	M2601-11	1	Solid	20.04	20	23.8	6.49	06/03/2021	17:32
11	M2601-11DUP	1	Solid	20.04	20	23.8	6.50	06/03/2021	17:34
12	CCV2	1	Water	NA	NA	20.4	12.02	06/03/2021	17:36

VB 114762

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2597 WorkList ID : 149472 Department : Wet-Chemistry Date : 06-03-2021 16:58:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/03/2021	Solid	M2601-03	pH	Cool 4 deg C	PSEG03	G11	RO-RBR200059	06/03/2021	9045D
06/03/2021	Solid	M2601-05	pH	Cool 4 deg C	PSEG03	G11	RO-CHRT-27166	06/03/2021	9045D
06/03/2021	Solid	M2601-07	pH	Cool 4 deg C	PSEG03	G11	RO-VNJ-222	06/03/2021	9045D
06/03/2021	Solid	M2601-09	pH	Cool 4 deg C	PSEG03	G11	RO-VNJ-212	06/03/2021	9045D
06/03/2021	Solid	M2601-11	pH	Cool 4 deg C	PSEG03	G11	RO-72-11929	06/03/2021	9045D

Date/Time 06/03/21 17:00
Raw Sample Received by: JA Cat (rem)
Raw Sample Relinquished by: KS (webchem)

Date/Time 06/03/21 18:00
Raw Sample Received by: KS (webchem)
Raw Sample Relinquished by: JA Cat (rem)