



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114763

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/03/2021	17:10
2	CAL2	1	Water	NA	NA	20.3	7.00	06/03/2021	17:11
3	CAL3	1	Water	NA	NA	20.3	10.02	06/03/2021	17:12
4	ICV	1	Water	NA	NA	20.3	6.99	06/03/2021	17:15
5	CCV1	1	Water	NA	NA	20.3	2.01	06/03/2021	17:16
6	M2609-01	1	Water	NA	NA	20.6	10.50	06/03/2021	17:40
7	M2609-01DUP	1	Water	NA	NA	20.7	10.51	06/03/2021	17:41
8	CCV2	1	Water	NA	NA	20.1	12.03	06/03/2021	17:45

LB 114763

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M2602 V WorkList ID : 149473 Department : Wet-Chemistry Date : 06-03-2021 16:59:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/03/2021	Water	M2609-01	Corrosivity	Cool 4 deg C	JACO05	F11	IDW-AQ-FT1-060321	06/03/2021	9040C

Date/Time 06/03/21 17:00
 Raw Sample Received by: JS (wet chem)
 Raw Sample Relinquished by: JS (wet chem)

Date/Time 06/03/21 18:00
 Raw Sample Received by: JS (wet chem)
 Raw Sample Relinquished by: JS (wet chem)