


LB114765

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.055	0.055	121	0.000	121					17
0.2	0.200	0.234	0.034	818	0.0 %	818					-6
2.5	2.500	2.362	-0.138	9131	0.0 %	9131					2
5.0	5.000	5.103	0.103	19833	0.0 %	19833					-2
7.5	7.500	7.377	-0.123	28717	0.0 %	28717					1
10.0	10.000	10.069	0.069	39228	0.0 %	39228					

LB114765 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 04 Jun 2021 09:25:58				Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units				
0.0 - 1	50	121	-	0.00007470A	PPB		04 Jun 2021 09:31:47	S	Std	
0.2 - 1	50.2	818	-	0.20007470A	PPB		04 Jun 2021 09:34:04	S	Std	
2.5 - 1	52.5	9131	-	2.50007470A	PPB		04 Jun 2021 09:36:21	S	Std	
5.0 - 1	53	19833	-	5.00007470A	PPB		04 Jun 2021 09:38:38	S	Std	
7.5 - 1	53.5	28717	-	7.50007470A	PPB		04 Jun 2021 09:40:54	S	Std	
10.0 - 1	54	39228	-	10.00007470A	PPB		04 Jun 2021 09:51:07	S	Std	
ICV47 - 1	ICV47	14599	3.7624	-7470A	PPB		04 Jun 2021 09:54:53	U	SMPL	
ICB47 - 1	ICB47	86	0.0461	-7470A	PPB		04 Jun 2021 09:57:07	U	SMPL	
CCV14 - 1	CCV14	19309	4.9684	-7470A	PPB		04 Jun 2021 09:59:23	U	SMPL	
CCB14 - 1	CCB14	71	0.0423	-7470A	PPB		04 Jun 2021 10:01:38	U	SMPL	
CRA - 1	CRA	744	0.2146	-7470A	PPB		04 Jun 2021 10:03:53	U	SMPL	
HighStd - 1	HighStd	38688	9.9307	-7470A	PPB		04 Jun 2021 10:06:08	U	SMPL	
ChkStd - 1	ChkStd	25950	6.6690	-7470A	PPB		04 Jun 2021 10:08:23	U	SMPL	
PB136816BL - 1	PBW	-126	-0.0082	-7470A	PPB		04 Jun 2021 10:10:39	U	SMPL	
PB136816BS - 1	LCSW	14349	3.6984	-7470A	PPB		04 Jun 2021 10:12:57	U	SMPL	
M2580-18 - 1	18-B1(0-2)	-74	0.0052	-7470A	PPB		04 Jun 2021 10:15:13	U	SMPL	
M2580-18DUP - 1	18-B1(0-2)DUP	14	0.0277	-7470A	PPB		04 Jun 2021 10:17:29	U	SMPL	
M2580-18MS - 1	18-B1(0-2)MS	13850	3.5706	-7470A	PPB		04 Jun 2021 10:19:45	U	SMPL	
M2580-18MSD - 1	18-B1(0-2)MSD	14156	3.6489	-7470A	PPB		04 Jun 2021 10:22:03	U	SMPL	
M2580-19 - 1	18-B1(2-4)	-49	0.0116	-7470A	PPB		04 Jun 2021 10:24:18	U	SMPL	
CCV15 - 1	CCV15	19055	4.9034	-7470A	PPB		04 Jun 2021 10:26:34	U	SMPL	
CCB15 - 1	CCB15	85	0.0459	-7470A	PPB		04 Jun 2021 10:28:48	U	SMPL	
M2580-20 - 1	18-B1(4-6)	23	0.0300	-7470A	PPB		04 Jun 2021 10:31:04	U	SMPL	
M2580-21 - 1	18-B2(0-2)	39	0.0341	-7470A	PPB		04 Jun 2021 10:33:19	U	SMPL	
M2580-22 - 1	18-B2(2-4)	52	0.0374	-7470A	PPB		04 Jun 2021 10:35:34	U	SMPL	
M2580-23 - 1	18-B2(4-6)	70	0.0420	-7470A	PPB		04 Jun 2021 10:37:50	U	SMPL	
M2580-24 - 1	18-B3(0-2)	50	0.0369	-7470A	PPB		04 Jun 2021 10:40:05	U	SMPL	
M2580-25 - 1	18-B3(2-4)	42	0.0349	-7470A	PPB		04 Jun 2021 10:42:21	U	SMPL	
M2580-26 - 1	18-B3(4-6)	47	0.0361	-7470A	PPB		04 Jun 2021 10:44:37	U	SMPL	
M2580-27 - 1	18-B4(0-2)	40	0.0343	-7470A	PPB		04 Jun 2021 10:46:54	U	SMPL	
M2580-28 - 1	18-B4(2-4)	38	0.0338	-7470A	PPB		04 Jun 2021 10:49:11	U	SMPL	
M2580-29 - 1	18-B4(4-6)	27	0.0310	-7470A	PPB		04 Jun 2021 10:51:29	U	SMPL	
CCV16 - 1	CCV16	18977	4.8834	-7470A	PPB		04 Jun 2021 10:53:44	U	SMPL	
CCB16 - 1	CCB16	62	0.0400	-7470A	PPB		04 Jun 2021 10:55:59	U	SMPL	
M2580-30 - 1	18-B5(0-2)	34	0.0328	-7470A	PPB		04 Jun 2021 10:58:14	U	SMPL	
M2580-31 - 1	18-B5(2-4)	82	0.0451	-7470A	PPB		04 Jun 2021 11:00:30	U	SMPL	
M2580-32 - 1	18-B5(4-6)	63	0.0402	-7470A	PPB		04 Jun 2021 11:02:45	U	SMPL	
M2580-33 - 1	18-B6(0-6)	57	0.0387	-7470A	PPB		04 Jun 2021 11:05:01	U	SMPL	
M2580-34 - 1	18-B6(6-12)	58	0.0390	-7470A	PPB		04 Jun 2021 11:07:17	U	SMPL	
PB136817BL - 1	PBW	49	0.0367	-7470A	PPB		04 Jun 2021 11:09:32	U	SMPL	
PB136817BS - 1	LCSW	14187	3.6569	-7470A	PPB		04 Jun 2021 11:11:49	U	SMPL	
M2583-19 - 1	33-34	378	0.1209	-7470A	PPB		04 Jun 2021 11:14:06	U	SMPL	
M2583-19DUP - 1	33-34DUP	413	0.1299	-7470A	PPB		04 Jun 2021 11:16:23	U	SMPL	
M2583-19MS - 1	33-34MS	15435	3.9765	-7470A	PPB		04 Jun 2021 11:18:38	U	SMPL	
CCV17 - 1	CCV17	19547	5.0294	-7470A	PPB		04 Jun 2021 11:20:54	U	SMPL	
CCB17 - 1	CCB17	57	0.0387	-7470A	PPB		04 Jun 2021 11:23:09	U	SMPL	
M2583-19MSD - 1	33-34MSD	15837	4.0794	-7470A	PPB		04 Jun 2021 11:25:25	U	SMPL	
M2583-20 - 1	35-38	1005	0.2815	-7470A	PPB		04 Jun 2021 11:27:41	U	SMPL	
M2583-21 - 1	39-41	912	0.2576	-7470A	PPB		04 Jun 2021 11:29:56	U	SMPL	
M2583-22 - 1	54-57	1880	0.5055	-7470A	PPB		04 Jun 2021 11:32:11	U	SMPL	
PB136816TB - 1	PB136779TB	10	0.0267	-7470A	PPB		04 Jun 2021 11:34:26	U	SMPL	
M2580-18LX5 - 1		93	0.0479	-7470A	PPB		04 Jun 2021 11:36:42	U	SMPL	
M2580-18A - 1		14955	3.8535	-7470A	PPB		04 Jun 2021 11:38:58	U	SMPL	
M2583-19LX5 - 1		104	0.0507	-7470A	PPB		04 Jun 2021 11:41:15	U	SMPL	
M2583-19A - 1		15595	4.0174	-7470A	PPB		04 Jun 2021 11:43:32	U	SMPL	
CCV18 - 1	CCV18	19364	4.9825	-7470A	PPB		04 Jun 2021 11:45:48	U	SMPL	
CCB18 - 1	CCB18	88	0.0466	-7470A	PPB		04 Jun 2021 11:48:03	U	SMPL	