



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114780

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/04/2021	09:05
2	M2583-19		1	8	100.00	>212.0	>212.0	06/04/2021	09:35
3	M2583-19DUP		1	8	100.00	>212.0	>212.0	06/04/2021	09:58
4	M2583-20		1	9	100.00	>212.0	>212.0	06/04/2021	10:25
5	M2583-21		1	9	100.00	>212.0	>212.0	06/04/2021	10:50
6	M2583-22		1	8	100.00	>212.0	>212.0	06/04/2021	11:15
7	M2609-01		1	9	100.00	>212.0	>212.0	06/04/2021	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114780

WorkList Name : fp6321

WorkList ID : 149463

Department : Wet-Chemistry

Date : 06-03-2021 12:41:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/17/2021	Water	M2609-01	Flash Point	Cool 4 deg C	JACO05	F11	IDW-AQ-FT1-060321	06/03/2021	1010A
06/07/2021	Water	M2583-19	Flash Point	Cool 4 deg C	PSEG03	G11	33-34	06/02/2021	1010A
06/07/2021	Water	M2583-20	Flash Point	Cool 4 deg C	PSEG03	G11	35-38	06/02/2021	1010A
06/07/2021	Water	M2583-21	Flash Point	Cool 4 deg C	PSEG03	G11	39-41	06/02/2021	1010A
06/07/2021	Water	M2583-22	Flash Point	Cool 4 deg C	PSEG03	G11	54-57	06/02/2021	1010A

Date/Time 06/04/21 08:30
 Raw Sample Received by: KS (Webster)
 Raw Sample Relinquished by: SP (Webster)

Date/Time 06/04/21 11:45
 Raw Sample Received by: SP (Webster)
 Raw Sample Relinquished by: KS (Webster)