



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114782

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/04/2021	12:30
2	CAL2	1	Water	NA	NA	20.4	7.01	06/04/2021	12:31
3	CAL3	1	Water	NA	NA	20.4	10.03	06/04/2021	12:32
4	ICV	1	Water	NA	NA	20.4	6.99	06/04/2021	12:35
5	CCV1	1	Water	NA	NA	20.6	2.01	06/04/2021	12:36
6	M2603-02	1	Solid	20.02	20	24.2	7.24	06/04/2021	12:39
7	M2603-02DUP	1	Solid	20.01	20	24.1	7.25	06/04/2021	12:40
8	M2603-06	1	Solid	20.03	20	24.1	9.87	06/04/2021	12:45
9	M2603-10	1	Solid	20.04	20	24.0	9.49	06/04/2021	12:47
10	M2603-14	1	Solid	20.02	20	23.5	8.96	06/04/2021	12:49
11	M2603-18	1	Solid	20.03	20	23.7	9.48	06/04/2021	12:52
12	M2603-22	1	Solid	20.04	20	24.0	10.60	06/04/2021	13:00
13	M2603-26	1	Solid	20.02	20	24.2	10.49	06/04/2021	13:02
14	CCV2	1	Water	NA	NA	20.3	12.02	06/04/2021	13:05

114782

WORKLIST(Hardcopy Internal Chain)

WorkList Name: CORROSIONITY M2603 WorkList ID: 149497 Department: Wet-Chemistry Date: 06-04-2021 12:07:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/02/2021	Solid	M2603-02	Corrosivity	Cool 4 deg C	ENTA04	F12	FAM4-WC1	06/02/2021	9045D
06/02/2021	Solid	M2603-06	Corrosivity	Cool 4 deg C	ENTA04	F12	FAM4-WC2	06/02/2021	9045D
06/02/2021	Solid	M2603-10	Corrosivity	Cool 4 deg C	ENTA04	F12	FAM4-WC3	06/02/2021	9045D
06/02/2021	Solid	M2603-14	Corrosivity	Cool 4 deg C	ENTA04	F12	FAM4-WC4	06/02/2021	9045D
06/02/2021	Solid	M2603-18	Corrosivity	Cool 4 deg C	ENTA04	F12	FAM4-WC5	06/02/2021	9045D
06/02/2021	Solid	M2603-22	Corrosivity	Cool 4 deg C	ENTA04	F12	FAM4-WC6	06/02/2021	9045D
06/02/2021	Solid	M2603-26	Corrosivity	Cool 4 deg C	ENTA04	F12	FAM4-WC7	06/02/2021	9045D

Date/Time 06/04/21 12:10
 Raw Sample Received by: JD Cut Chem
 Raw Sample Relinquished by: [Signature]

Date/Time 06/04/21 16:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: JD Cut Chem