



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114792

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	06/04/2021	15:37
2	CAL2	1	Water	NA	NA	20.6	7.00	06/04/2021	15:38
3	CAL3	1	Water	NA	NA	20.6	10.02	06/04/2021	15:39
4	ICV	1	Water	NA	NA	20.6	7.00	06/04/2021	15:40
5	CCV1	1	Water	NA	NA	20.6	2.01	06/04/2021	15:42
6	M2615-01	1	Water	NA	NA	24.5	7.68	06/04/2021	15:45
7	M2615-01DUP	1	Water	NA	NA	20.6	7.69	06/04/2021	15:46
8	M2617-01	1	Water	NA	NA	23.4	5.01	06/04/2021	15:48
9	CCV2	1	Water	NA	NA	20.6	12.02	06/04/2021	15:50

WORKLIST(Hardcopy Internal Chain)

103 114792

WorkList Name : PH W M2617 WorkList ID : 149511 Department : Wet-Chemistry Date : 06-04-2021 14:46:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/04/2021	Water	M2615-01	pH	Cool 4 deg C	PSEG03	G12	1245	06/04/2021	9040C
06/03/2021	Water	M2617-01	pH	Cool 4 deg C	STER08	F21	POOL-WATER	06/03/2021	9040C

Date/Time 06/04/21 15:30
 Raw Sample Received by: JA 104 120m
 Raw Sample Relinquished by: CL SM

Date/Time 06/04/21 16:30
 Raw Sample Received by: CL SM
 Raw Sample Relinquished by: JA 104 120m