



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 06/05/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 06/04/2021 17:10 DATE & TIME OUT1: 06/05/2021 08:15OVEN IN TEMP2: 103 °COVEN OUT TEMP2: 103 °CDATE & TIME IN2: 06/05/2021 09:00 DATE & TIME OUT2: 06/05/2021 10:30Run Number: LB114793BalanceID: WC SC-3OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB114793BL	LB114793BL	68.5417	68.5417	100	68.5418	68.5418	68.5418	0.0001	1
2	M2615-01	1245	75.5534	75.5534	100	75.7376	75.7376	75.7376	0.1842	1842
3	M2615-01DUP	1245DUP	68.8202	68.8202	100	68.9984	68.9984	68.9984	0.1782	1782

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

114793

WorkList Name : ts w m2615 WorkList ID : 149516 Department : Wet-Chemistry Date : 06-04-2021 15:51:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/09/2021	Water	M2615-01	TS	Cool 4 deg C	PSEG03	G12	1245	06/04/2021	SM2540 B

Date/Time 06/04/21 11:00
 Raw Sample Received by: go 1st 12pm
 Raw Sample Relinquished by: CP SM

Date/Time 06/04/21 17:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: go 1st 6pm