



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 06/04/2021 17:10 TEMP1 OUT: 104 °C 06/05/2021 08:15
TEMP2 IN: 103 °C 06/05/2021 09:00 TEMP2 OUT: 103 °C 06/05/2021 10:30
TEMP3 IN: 540 °C 06/05/2021 11:00 TEMP3 OUT: 540 °C 06/05/2021 12:30
TEMP4 IN: 560 °C 06/05/2021 13:00 TEMP4 OUT: 560 °C 06/05/2021 14:30

Run Number: LB114794

SUPERVISOR: apatel

ANALYST: jignesh

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB114794BL	68.5417	68.5417	100	68.5418	68.5418	68.5418	68.5418	0.0000	0
2	M2615-01	75.5534	75.5534	100	75.7376	75.7376	75.7009	75.7009	0.0367	367
3	M2615-01DUP	68.8202	68.8202	100	68.9984	68.9984	68.9626	68.9626	0.0358	358

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \times 1000000$

WORKLIST(Hardcopy Internal Chain)

RP 114794

WorkList Name : tvs w m2615

WorkList ID : 149517

Department : Wet-Chemistry

Date : 06-04-2021 15:51:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/09/2021	Water	M2615-01	TVS	Cool 4 deg C	PSEG03	G12	1245	06/04/2021	SM2540 G

Date/Time 06/04/21 16:00
 Raw Sample Received by: JO 11/04/21 Dem
 Raw Sample Relinquished by: CP SM

Date/Time 06/04/21 17:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JO CP Dem