



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114805

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	06/05/2021	13:58
2	M2615-01		1	7	100.00	>212.0	>212.0	06/05/2021	14:25
3	M2615-01DUP		1	9	100.00	>212.0	>212.0	06/05/2021	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114805

WorkList Name : fp6421 WorkList ID : 149529 Department : Wet-Chemistry Date : 06-04-2021 17:36:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/09/2021	Water	M2615-01	R Flash Point	Cool 4 deg C	PSEG03	G12	1245	06/04/2021	1010A

Date/Time 06/05/21 13:30
 Raw Sample Received by: KS (Wehner)
 Raw Sample Relinquished by: RJ C EKF-1006

Date/Time 06/05/21 14:50
 Raw Sample Received by: RJ C EKF-1006
 Raw Sample Relinquished by: KS (Wehner)