

LB 114815


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

3852E



A= 0.0000e+000

B= 2.5884e-004

C= -9.9629e-003

Rho= 0.9999400

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.013	0.013	87	0.000	87					-
0.2	0.200	0.190	-0.010	772	0.0 %	772					-5
2.5	2.500	2.445	-0.055	9485	0.0 %	9485					-2
5.0	5.000	5.068	0.068	19617	0.0 %	19617					1
7.5	7.500	7.523	0.023	29102	0.0 %	29102					0
10.0	10.000	9.962	-0.038	38526	0.0 %	38526					0

LB114815 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Jun 2021 09:20:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	87	-	0.00007470A	PPB	07 Jun 2021 09:36:16	S	Std	
0.2 - 1	772	-	0.20007470A	PPB	07 Jun 2021 09:38:33	S	Std	
2.5 - 1	9485	-	2.50007470A	PPB	07 Jun 2021 09:40:50	S	Std	
5.0 - 1	19617	-	5.00007470A	PPB	07 Jun 2021 09:43:07	S	Std	
7.5 - 1	29102	-	7.50007470A	PPB	07 Jun 2021 09:45:23	S	Std	
10.0 - 1	38526	-	10.00007470A	PPB	07 Jun 2021 09:50:13	S	Std	
ICV49 - 1	ICV49	4.1408	-7470A	PPB	07 Jun 2021 09:52:49	U	SMPL	
ICB49 - 1	ICB49	-18	-0.0146	-7470A	PPB	07 Jun 2021 09:55:03	U	SMPL
CCV26 - 1	CCV26	19388	5.0084	-7470A	PPB	07 Jun 2021 09:57:18	U	SMPL
CCB26 - 1	CCB26	-20	-0.0151	-7470A	PPB	07 Jun 2021 09:59:33	U	SMPL
CRA - 1	CRA	759	0.1865	-7470A	PPB	07 Jun 2021 10:01:48	U	SMPL
HighStd - 1	HighStd	38260	9.8932	-7470A	PPB	07 Jun 2021 10:04:03	U	SMPL
ChkStd - 1	ChkStd	27449	7.0949	-7470A	PPB	07 Jun 2021 10:06:18	U	SMPL
PB136839BL - 1	PBW	71	0.0084	-7470A	PPB	07 Jun 2021 10:08:34	U	SMPL
PB136839BS - 1	LCSW	15525	4.0085	-7470A	PPB	07 Jun 2021 10:10:51	U	SMPL
M2589-21 - 1	18-B7(0-2)	78	0.0102	-7470A	PPB	07 Jun 2021 10:13:07	U	SMPL
M2589-21DUP - 1	18-B7(0-2)DUP	169	0.0338	-7470A	PPB	07 Jun 2021 10:15:24	U	SMPL
M2589-21MS - 1	18-B7(0-2)MS	15105	3.8998	-7470A	PPB	07 Jun 2021 10:17:41	U	SMPL
M2589-21MSD - 1	18-B7(0-2)MSD	15619	4.0328	-7470A	PPB	07 Jun 2021 10:19:58	U	SMPL
M2589-22 - 1	18-B7(2-4)	82	0.0113	-7470A	PPB	07 Jun 2021 10:22:13	U	SMPL
CCV27 - 1	CCV27	19335	4.9947	-7470A	PPB	07 Jun 2021 10:24:29	U	SMPL
CCB27 - 1	CCB27	13	-0.0066	-7470A	PPB	07 Jun 2021 10:26:45	U	SMPL
M2589-23 - 1	18-B7(4-6)	92	0.0139	-7470A	PPB	07 Jun 2021 10:29:00	U	SMPL
M2589-24 - 1	18-B8(0-2)	89	0.0131	-7470A	PPB	07 Jun 2021 10:31:15	U	SMPL
M2589-25 - 1	18-B8(2-4)	48	0.0025	-7470A	PPB	07 Jun 2021 10:33:31	U	SMPL
M2589-26 - 1	18-B8(4-6)	51	0.0032	-7470A	PPB	07 Jun 2021 10:35:46	U	SMPL
M2589-27 - 1	18-B9(0-2)	236	0.0511	-7470A	PPB	07 Jun 2021 10:38:02	U	SMPL
M2589-28 - 1	18-B9(2-4)	247	0.0540	-7470A	PPB	07 Jun 2021 10:40:18	U	SMPL
M2589-29 - 1	18-B9(4-6)	267	0.0591	-7470A	PPB	07 Jun 2021 10:42:35	U	SMPL
M2589-30 - 1	FOUNDATION-EXCAVATION-SAMPLE-A	283	0.0633	-7470A	PPB	07 Jun 2021 10:44:51	U	SMPL
M2589-31 - 1	18-B10(0-2)	246	0.0537	-7470A	PPB	07 Jun 2021 10:47:08	U	SMPL
M2589-32 - 1	18-B10(2-4)	209	0.0441	-7470A	PPB	07 Jun 2021 10:49:23	U	SMPL
CCV28 - 1	CCV28	19394	5.0099	-7470A	PPB	07 Jun 2021 10:51:39	U	SMPL
CCB28 - 1	CCB28	13	-0.0066	-7470A	PPB	07 Jun 2021 10:53:54	U	SMPL
M2589-33 - 1	18-B10(4-6)	63	0.0063	-7470A	PPB	07 Jun 2021 10:56:09	U	SMPL
M2589-34 - 1	18-B11(0-2)	71	0.0084	-7470A	PPB	07 Jun 2021 10:58:25	U	SMPL
M2589-35 - 1	18-B11(2-4)	99	0.0157	-7470A	PPB	07 Jun 2021 11:00:40	U	SMPL
M2589-36 - 1	18-B11(4-6)	83	0.0115	-7470A	PPB	07 Jun 2021 11:02:55	U	SMPL
M2589-37 - 1	FOUNDATION-EXCAVATION-SAMPLE-B	383	0.0892	-7470A	PPB	07 Jun 2021 11:05:11	U	SMPL
M2589-38 - 1	18-B12(0-2)	63	0.0063	-7470A	PPB	07 Jun 2021 11:07:27	U	SMPL
M2589-39 - 1	18-B12(2-4)	104	0.0170	-7470A	PPB	07 Jun 2021 11:09:43	U	SMPL
M2589-40 - 1	18-B12(4-6)	92	0.0139	-7470A	PPB	07 Jun 2021 11:12:00	U	SMPL
PB136842BL - 1	PBW	25	-0.0035	-7470A	PPB	07 Jun 2021 11:17:05	U	SMPL
PB136842BS - 1	LCSW	15872	4.0983	-7470A	PPB	07 Jun 2021 11:19:21	U	SMPL
CCV29 - 1	CCV29	19212	4.9628	-7470A	PPB	07 Jun 2021 11:21:36	U	SMPL
CCB29 - 1	CCB29	-1	-0.0102	-7470A	PPB	07 Jun 2021 11:23:51	U	SMPL
M2262-07 - 1	LOD-MDL-WATER-SOIL-01-QT2-2021	256	0.0563	-7470A	PPB	07 Jun 2021 11:29:01	U	SMPL
M2262-08 - 1	LOQ-WATER-02-QT2-2021	763	0.1875	-7470A	PPB	07 Jun 2021 11:33:50	U	SMPL
M2609-01 - 1	IDW-AQ-FT1-060321	135	0.0250	-7470A	PPB	07 Jun 2021 11:36:05	U	SMPL
M2615-01 - 1	1245	162	0.0320	-7470A	PPB	07 Jun 2021 11:38:20	U	SMPL
M2615-01DUP - 1	1245DUP	205	0.0431	-7470A	PPB	07 Jun 2021 11:40:36	U	SMPL
M2615-01MS - 1	1245MS	15894	4.1040	-7470A	PPB	07 Jun 2021 11:42:52	U	SMPL
M2615-01MSD - 1	1245MSD	15912	4.1087	-7470A	PPB	07 Jun 2021 11:45:09	U	SMPL
CCV30 - 1	CCV30	19430	5.0193	-7470A	PPB	07 Jun 2021 11:47:26	U	SMPL
CCB30 - 1	CCB30	-6	-0.0115	-7470A	PPB	07 Jun 2021 11:49:43	U	SMPL
PB136839TB - 1	PB136805TB	88	0.0128	-7470A	PPB	07 Jun 2021 11:51:59	U	SMPL
M2589-21LX5 - 1		158	0.0309	-7470A	PPB	07 Jun 2021 11:54:14	U	SMPL
M2589-21A - 1		16042	4.1423	-7470A	PPB	07 Jun 2021 11:56:30	U	SMPL
M2615-01LX5 - 1		57	0.0048	-7470A	PPB	07 Jun 2021 11:58:45	U	SMPL
M2615-01A - 1		15677	4.0478	-7470A	PPB	07 Jun 2021 12:01:01	U	SMPL
CCV31 - 1	CCV31	19391	5.0092	-7470A	PPB	07 Jun 2021 12:03:16	U	SMPL
CCB31 - 1	CCB31	-19	-0.0149	-7470A	PPB	07 Jun 2021 12:05:31	U	SMPL