



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114834

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	06/07/2021	17:00
2	CAL2	1	Water	NA	NA	21.4	7.01	06/07/2021	17:01
3	CAL3	1	Water	NA	NA	21.4	10.04	06/07/2021	17:02
4	ICV	1	Water	NA	NA	21.4	7.00	06/07/2021	17:05
5	CCV1	1	Water	NA	NA	21.5	2.01	06/07/2021	17:06
6	M2627-01	1	Water	NA	NA	25.5	5.50	06/07/2021	17:10
7	M2627-01DUP	1	Water	NA	NA	25.6	5.51	06/07/2021	17:11
8	M2627-03	1	Water	NA	NA	25.9	6.42	06/07/2021	17:12
9	CCV2	1	Water	NA	NA	21.4	12.02	06/07/2021	17:15

10114834
10114834

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2634 WorkList ID : 149605 Department : Wet-Chemistry Date : 06-07-2021 16:40:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/07/2021	Water	M2627-01	pH	Cool 4 deg C	PSEG03	G21	33226	06/07/2021	9040C
06/07/2021	Water	M2627-03	pH	Cool 4 deg C	PSEG03	G21	33227	06/07/2021	9040C

Date/Time 06/07/21 16:45
Raw Sample Received by: JO (Wetchem)
Raw Sample Relinquished by: KS (Wetchem)

Date/Time 06/07/21 17:35
Raw Sample Received by: KS (Wetchem)
Raw Sample Relinquished by: JO (Wetchem)