

LB114836


7470A

INSTRUMENT 110: CV1

Linear

 μ Abs.:

38928



A= 0.0000e+000

B= 2.5809e-004

C= -1.9275e-002

Rho= 0.9996109

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	39	0.000	39				
0.2	0.200	0.182	-0.018	778	0.0 %	778				
2.5	2.500	2.473	-0.027	9657	0.0 %	9657				
5.0	5.000	5.188	0.188	20178	0.0 %	20178				
7.5	7.500	7.338	-0.162	28508	0.0 %	28508				
10.0	10.000	10.028	0.028	38928	0.0 %	38928				

0/00
 1
 -9
 -1
 4
 -2
 0

LB114836 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Jun 2021 10:16:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	7470A	PPB	08 Jun 2021 10:17:47	S	Std
0.2 - 1	50.2	778	-	0.2000	7470A	PPB	08 Jun 2021 10:20:03	S	Std
2.5 - 1	52.5	9657	-	2.5000	7470A	PPB	08 Jun 2021 10:22:20	S	Std
5.0 - 1	55	20178	-	5.0000	7470A	PPB	08 Jun 2021 10:24:36	S	Std
7.5 - 1	57.5	28508	-	7.5000	7470A	PPB	08 Jun 2021 10:26:53	S	Std
10.0 - 1	511	38928	-	10.0000	7470A	PPB	08 Jun 2021 10:36:51	S	Std
ICV52 - 1	ICV52	15684	4.0286	-	7470A	PPB	08 Jun 2021 10:39:31	U	SMPL
ICB52 - 1	ICB52	-69	-0.0371	-	7470A	PPB	08 Jun 2021 10:41:46	U	SMPL
CCV42 - 1	CCV42	18087	4.6488	-	7470A	PPB	08 Jun 2021 10:44:01	U	SMPL
CCB42 - 1	CCB42	-46	-0.0311	-	7470A	PPB	08 Jun 2021 10:46:16	U	SMPL
CRA - 1	CRA	767	0.1787	-	7470A	PPB	08 Jun 2021 10:48:30	U	SMPL
HighStd - 1	HighStd	39431	10.1575	-	7470A	PPB	08 Jun 2021 10:50:45	U	SMPL
ChkStd - 1	ChkStd	26645	6.8576	-	7470A	PPB	08 Jun 2021 10:53:01	U	SMPL
PB136890BL - 1	PBW	-57	-0.0340	-	7470A	PPB	08 Jun 2021 10:55:17	U	SMPL
PB136890BS - 1	LCSW	13779	3.5370	-	7470A	PPB	08 Jun 2021 10:57:32	U	SMPL
M2597-02 - 1	124019	-33	-0.0278	-	7470A	PPB	08 Jun 2021 10:59:48	U	SMPL
M2597-04 - 1	124042	33	-0.0108	-	7470A	PPB	08 Jun 2021 11:02:04	U	SMPL
M2599-03 - 1	339	33	-0.0108	-	7470A	PPB	08 Jun 2021 11:04:21	U	SMPL
M2603-02 - 1	FAM4-WC1	45	-0.0077	-	7470A	PPB	08 Jun 2021 11:06:38	U	SMPL
M2603-06 - 1	FAM4-WC2	43	-0.0082	-	7470A	PPB	08 Jun 2021 11:08:53	U	SMPL
CCV43 - 1	CCV43	18092	4.6501	-	7470A	PPB	08 Jun 2021 11:11:08	U	SMPL
CCB43 - 1	CCB43	-58	-0.0342	-	7470A	PPB	08 Jun 2021 11:13:23	U	SMPL
M2603-10 - 1	FAM4-WC3	60	-0.0038	-	7470A	PPB	08 Jun 2021 11:15:38	U	SMPL
M2603-14 - 1	FAM4-WC4	21	-0.0139	-	7470A	PPB	08 Jun 2021 11:17:53	U	SMPL
M2603-18 - 1	FAM4-WC5	45	-0.0077	-	7470A	PPB	08 Jun 2021 11:20:08	U	SMPL
M2603-22 - 1	FAM4-WC6	29	-0.0118	-	7470A	PPB	08 Jun 2021 11:22:24	U	SMPL
M2603-26 - 1	FAM4-WC7	31	-0.0113	-	7470A	PPB	08 Jun 2021 11:24:40	U	SMPL
M2616-03 - 1	BSS-SOIL-PILE	35	-0.0102	-	7470A	PPB	08 Jun 2021 11:26:56	U	SMPL
M2616-03DUP - 1	BSS-SOIL-PILEDUP	37	-0.0097	-	7470A	PPB	08 Jun 2021 11:29:12	U	SMPL
M2616-03MS - 1	BSS-SOIL-PILEMS	13060	3.3514	-	7470A	PPB	08 Jun 2021 11:31:29	U	SMPL
M2616-03MSD - 1	BSS-SOIL-PILEMSD	12501	3.2071	-	7470A	PPB	08 Jun 2021 11:33:46	U	SMPL
PB136891BL - 1	PBW	-39	-0.0293	-	7470A	PPB	08 Jun 2021 11:36:01	U	SMPL
CCV44 - 1	CCV44	18219	4.6829	-	7470A	PPB	08 Jun 2021 11:38:17	U	SMPL
CCB44 - 1	CCB44	-41	-0.0299	-	7470A	PPB	08 Jun 2021 11:40:32	U	SMPL
PB136891BS - 1	LCSW	13702	3.5171	-	7470A	PPB	08 Jun 2021 11:42:48	U	SMPL
M2612-21 - 1	18-B12(6-10)	-36	-0.0286	-	7470A	PPB	08 Jun 2021 11:45:03	U	SMPL
M2612-21DUP - 1	18-B12(6-10)DUP	34	-0.0105	-	7470A	PPB	08 Jun 2021 11:50:39	U	SMPL
M2612-21MS - 1	18-B12(6-10)MS	13627	3.4977	-	7470A	PPB	08 Jun 2021 11:52:55	U	SMPL
M2612-21MSD - 1	18-B12(6-10)MSD	13361	3.4291	-	7470A	PPB	08 Jun 2021 11:55:10	U	SMPL
M2612-22 - 1	18-B12(10-14)	-46	-0.0311	-	7470A	PPB	08 Jun 2021 11:57:26	U	SMPL
M2612-23 - 1	18-B12(14-18)	22	-0.0136	-	7470A	PPB	08 Jun 2021 11:59:43	U	SMPL
M2612-24 - 1	18-B12(18-22)	30	-0.0115	-	7470A	PPB	08 Jun 2021 12:01:59	U	SMPL
M2612-25 - 1	18-B12(22-26)	23	-0.0133	-	7470A	PPB	08 Jun 2021 12:04:16	U	SMPL
M2612-26 - 1	18-B12(26-30)	50	-0.0064	-	7470A	PPB	08 Jun 2021 12:06:31	U	SMPL
CCV45 - 1	CCV45	18640	4.7915	-	7470A	PPB	08 Jun 2021 12:08:47	U	SMPL
CCB45 - 1	CCB45	-45	-0.0309	-	7470A	PPB	08 Jun 2021 12:11:02	U	SMPL
M2612-27 - 1	18-B12(30-34)	31	-0.0113	-	7470A	PPB	08 Jun 2021 12:13:17	U	SMPL
M2612-28 - 1	18-B12(34-38)	32	-0.0110	-	7470A	PPB	08 Jun 2021 12:15:33	U	SMPL
M2612-29 - 1	18-B12(38-42)	21	-0.0139	-	7470A	PPB	08 Jun 2021 12:17:48	U	SMPL
M2612-30 - 1	18-B12(42-46)	17	-0.0149	-	7470A	PPB	08 Jun 2021 12:20:03	U	SMPL
M2612-31 - 1	18-B12(46-50)	49	-0.0066	-	7470A	PPB	08 Jun 2021 12:22:19	U	SMPL
M2612-32 - 1	18-B12(50-54)	25	-0.0128	-	7470A	PPB	08 Jun 2021 12:24:35	U	SMPL
M2612-33 - 1	18-B12(54-58)	44	-0.0079	-	7470A	PPB	08 Jun 2021 12:26:51	U	SMPL
M2612-34 - 1	18-B12(58-62)	34	-0.0105	-	7470A	PPB	08 Jun 2021 12:29:08	U	SMPL
M2612-35 - 1	18-B12(62-66)	32	-0.0110	-	7470A	PPB	08 Jun 2021 12:31:25	U	SMPL
M2612-36 - 1	18-B12(66-70)	46	-0.0074	-	7470A	PPB	08 Jun 2021 12:33:41	U	SMPL
CCV46 - 1	CCV46	18628	4.7884	-	7470A	PPB	08 Jun 2021 12:35:57	U	SMPL
CCB46 - 1	CCB46	-42	-0.0301	-	7470A	PPB	08 Jun 2021 12:38:12	U	SMPL
M2612-37 - 1	18-B12(70-74)	42	-0.0084	-	7470A	PPB	08 Jun 2021 12:40:27	U	SMPL
M2612-38 - 1	18-B12(74-78)	44	-0.0079	-	7470A	PPB	08 Jun 2021 12:42:43	U	SMPL
M2612-39 - 1	18-B12(78-82)	59	-0.0040	-	7470A	PPB	08 Jun 2021 12:44:58	U	SMPL
M2612-40 - 1	18-B12(82-86)	54	-0.0053	-	7470A	PPB	08 Jun 2021 12:47:13	U	SMPL
PB136903BL - 1	PBW	51	-0.0061	-	7470A	PPB	08 Jun 2021 12:49:30	U	SMPL
PB136903BS - 1	LCSW	15473	3.9742	-	7470A	PPB	08 Jun 2021 12:54:41	U	SMPL
M2613-14 - 1	18-B12(86-90)	0	-0.0193	-	7470A	PPB	08 Jun 2021 12:56:57	U	SMPL
M2613-14DUP - 1	18-B12(86-90)DUP	39	-0.0092	-	7470A	PPB	08 Jun 2021 12:59:14	U	SMPL
M2613-14MS - 1	18-B12(86-90)MS	14025	3.6005	-	7470A	PPB	08 Jun 2021 13:01:30	U	SMPL
M2613-14MSD - 1	18-B12(86-90)MSD	13630	3.4985	-	7470A	PPB	08 Jun 2021 13:03:46	U	SMPL
CCV47 - 1	CCV47	18322	4.7095	-	7470A	PPB	08 Jun 2021 13:06:03	U	SMPL
CCB47 - 1	CCB47	-66	-0.0363	-	7470A	PPB	08 Jun 2021 13:08:19	U	SMPL
M2613-15 - 1	18-B12(90-94)	45	-0.0077	-	7470A	PPB	08 Jun 2021 13:10:34	U	SMPL
M2613-16 - 1	18-B12(94-98)	42	-0.0084	-	7470A	PPB	08 Jun 2021 13:12:50	U	SMPL
M2613-17 - 1	18-B12(98-102)	24	-0.0131	-	7470A	PPB	08 Jun 2021 13:15:06	U	SMPL
M2613-18 - 1	18-B12(102-106)	40	-0.0090	-	7470A	PPB	08 Jun 2021 13:17:22	U	SMPL
M2613-19 - 1	18-B12(106-110)	32	-0.0110	-	7470A	PPB	08 Jun 2021 13:19:38	U	SMPL
M2613-20 - 1	18-B12(110-114)	38	-0.0095	-	7470A	PPB	08 Jun 2021 13:21:54	U	SMPL
M2613-21 - 1	18-B12(114-118)	65	-0.0025	-	7470A	PPB	08 Jun 2021 13:24:11	U	SMPL
M2613-22 - 1	18-B12(118-122)	56	-0.0048	-	7470A	PPB	08 Jun 2021 13:26:27	U	SMPL
M2613-23 - 1	18-B12(122-126)	51	-0.0061	-	7470A	PPB	08 Jun 2021 13:28:45	U	SMPL
M2613-24 - 1	18-B12(126-130)	29	-0.0118	-	7470A	PPB	08 Jun 2021 13:31:00	U	SMPL
CCV48 - 1	CCV48	18190	4.6754	-	7470A	PPB	08 Jun 2021 13:33:16	U	SMPL

LB114836 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Jun 2021 10:16:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB48 - 1	CCB48	-45	-0.0309		-7470A	PPB	08 Jun 2021 13:35:32	U	SMPL
M2613-25 - 1	18-B12(130-134)	28	-0.0120		-7470A	PPB	08 Jun 2021 13:37:48	U	SMPL
M2613-26 - 1	18-B12(134-138)	16	-0.0151		-7470A	PPB	08 Jun 2021 13:40:04	U	SMPL
PB136890TB - 1	PB136828TB	45	-0.0077		-7470A	PPB	08 Jun 2021 13:42:21	U	SMPL
PB136891TB - 1	PB136856TB	56	-0.0048		-7470A	PPB	08 Jun 2021 13:44:38	U	SMPL
PB136903TB - 1	PB136877TB	56	-0.0048		-7470A	PPB	08 Jun 2021 13:46:54	U	SMPL
CCV49 - 1	CCV49	18476	4.7492		-7470A	PPB	08 Jun 2021 13:49:11	U	SMPL
CCB49 - 1	CCB49	-54	-0.0332		-7470A	PPB	08 Jun 2021 13:51:27	U	SMPL
M2616-03LX5 - 1		35	-0.0102		-7470A	PPB	08 Jun 2021 13:53:44	U	SMPL
M2616-03A - 1		15398	3.9548		-7470A	PPB	08 Jun 2021 13:56:01	U	SMPL
M2612-21LX5 - 1		-51	-0.0324		-7470A	PPB	08 Jun 2021 13:58:17	U	SMPL
M2612-21A - 1		15510	3.9837		-7470A	PPB	08 Jun 2021 14:00:33	U	SMPL
M2613-14LX5 - 1		-28	-0.0265		-7470A	PPB	08 Jun 2021 14:02:50	U	SMPL
M2613-14A - 1		14984	3.8480		-7470A	PPB	08 Jun 2021 14:05:06	U	SMPL
CCV50 - 1	CCV50	18090	4.6496		-7470A	PPB	08 Jun 2021 14:07:22	U	SMPL
CCB50 - 1	CCB50	-63	-0.0355		-7470A	PPB	08 Jun 2021 14:09:39	U	SMPL