



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114847

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	06/08/2021	15:20
2	M2627-01		1	9	100.00	>212.0	>212.0	06/08/2021	15:45
3	M2627-01DUP		1	9	100.00	>212.0	>212.0	06/08/2021	16:10
4	M2627-03		1	8	100.00	>212.0	>212.0	06/08/2021	16:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114847

WorkList Name : fp6521 WorkList ID : 149556 Department : Wet-Chemistry Date : 06-05-2021 16:25:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/14/2021	Water	M2627-01	C Flash Point	Cool 4 deg C	PSEG03	G21	33226	06/07/2021	1010A
06/14/2021	Water	M2627-03	C Flash Point	Cool 4 deg C	PSEG03	G21	33227	06/07/2021	1010A

Date/Time 06/08/21 14:50
 Raw Sample Received by: KS (Webster)
 Raw Sample Relinquished by: CP SM

Date/Time 06/08/21 16:40
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: KS (Webster)