

LB114851


7471B

INSTRUMENT ID: CV2

Linear

μ Abs.:

50555



A= 0.0000e+000

B= 1.9838e-004

C= 3.7628e-002

Rho= 0.9999027

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.040	0.040	14	0.000	14					13
0.2	0.200	0.225	0.025	942	0.0 %	942					-2
2.5	2.500	2.438	-0.062	12102	0.0 %	12102					0
5.0	5.000	5.003	0.003	25032	0.0 %	25032					-1
7.5	7.500	7.426	-0.074	37246	0.0 %	37246					1
10.0	10.000	10.067	0.067	50555	0.0 %	50555					

LB114851 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 08 Jun 2021 15:28:55					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	58	14	-	0.0000	7471B	PPB	08 Jun 2021 15:30:54	S	Std
0.2 - 1	50.2	942	-	0.2000	7471B	PPB	08 Jun 2021 15:33:11	S	Std
2.5 - 1	52.5	12102	-	2.5000	7471B	PPB	08 Jun 2021 15:35:28	S	Std
5.0 - 1	55	25032	-	5.0000	7471B	PPB	08 Jun 2021 15:37:45	S	Std
7.5 - 1	57.5	37246	-	7.5000	7471B	PPB	08 Jun 2021 15:40:01	S	Std
10.0 - 1	51.5	50555	-	10.0000	7471B	PPB	08 Jun 2021 15:42:17	S	Std
ICV53 - 1	ICV53	19971	3.9995	-	7471B	PPB	08 Jun 2021 15:44:58	U	SMPL
ICB53 - 1	ICB53	-92	0.0194	-	7471B	PPB	08 Jun 2021 15:47:13	U	SMPL
CCV51 - 1	CCV51	24240	4.8464	-	7471B	PPB	08 Jun 2021 15:52:02	U	SMPL
CCB51 - 1	CCB51	-126	0.0126	-	7471B	PPB	08 Jun 2021 15:54:17	U	SMPL
CRA - 1	CRA	896	0.2154	-	7471B	PPB	08 Jun 2021 15:56:33	U	SMPL
HighStd - 1	HighStd	50118	9.9800	-	7471B	PPB	08 Jun 2021 15:58:48	U	SMPL
ChkStd - 1	ChkStd	34349	6.8518	-	7471B	PPB	08 Jun 2021 16:01:02	U	SMPL
PB136915BL - 1	PBS	-153	0.0073	-	7471B	PPB	08 Jun 2021 16:03:19	U	SMPL
PB136915BS - 1	LCSS	20176	4.0401	-	7471B	PPB	08 Jun 2021 16:05:36	U	SMPL
M2625-01 - 1	18-B12(138-142)	270	0.0912	-	7471B	PPB	08 Jun 2021 16:12:41	U	SMPL
M2625-01DUP - 1	18-B12(138-142)DUP	306	0.0983	-	7471B	PPB	08 Jun 2021 16:16:02	U	SMPL
M2625-01MS - 1	18-B12(138-142)MS	23459	4.6914	-	7471B	PPB	08 Jun 2021 16:18:18	U	SMPL
M2625-01MSD - 1	18-B12(138-142)MSD	19928	3.9909	-	7471B	PPB	08 Jun 2021 16:20:35	U	SMPL
M2625-02 - 1	18-B12(142-146)	189	0.0751	-	7471B	PPB	08 Jun 2021 16:22:50	U	SMPL
CCV52 - 1	CCV52	26735	5.3413	-	7471B	PPB	08 Jun 2021 16:25:06	U	SMPL
CCB52 - 1	CCB52	-86	0.0206	-	7471B	PPB	08 Jun 2021 16:27:21	U	SMPL
M2625-03 - 1	18-B12(146-150)	745	0.1854	-	7471B	PPB	08 Jun 2021 16:29:36	U	SMPL
M2625-04 - 1	18-B12(150-154)	637	0.1640	-	7471B	PPB	08 Jun 2021 16:31:52	U	SMPL
M2625-05 - 1	18-B12(154-158)	664	0.1694	-	7471B	PPB	08 Jun 2021 16:34:07	U	SMPL
M2625-06 - 1	18-B12(158-162)	474	0.1317	-	7471B	PPB	08 Jun 2021 16:36:22	U	SMPL
M2625-07 - 1	18-B12(162-166)	230	0.0833	-	7471B	PPB	08 Jun 2021 16:38:38	U	SMPL
M2625-08 - 1	18-B12(166-170)	438	0.1245	-	7471B	PPB	08 Jun 2021 16:40:54	U	SMPL
M2625-09 - 1	18-B12(170-174)	1394	0.3142	-	7471B	PPB	08 Jun 2021 16:43:10	U	SMPL
M2625-10 - 1	18-B12(174-178)	838	0.2039	-	7471B	PPB	08 Jun 2021 16:45:26	U	SMPL
M2625-11 - 1	18-B12(178-182)	568	0.1503	-	7471B	PPB	08 Jun 2021 16:47:43	U	SMPL
M2625-12 - 1	18-B12(182-186)	569	0.1505	-	7471B	PPB	08 Jun 2021 16:49:58	U	SMPL
CCV53 - 1	CCV53	26973	5.3885	-	7471B	PPB	08 Jun 2021 16:52:14	U	SMPL
CCB53 - 1	CCB53	-65	0.0247	-	7471B	PPB	08 Jun 2021 16:54:29	U	SMPL
M2625-13 - 1	18-B12(186-190)	936	0.2233	-	7471B	PPB	08 Jun 2021 16:56:44	U	SMPL
M2625-14 - 1	18-B12(190-194)	49	0.0473	-	7471B	PPB	08 Jun 2021 16:58:59	U	SMPL
M2625-15 - 1	18-B12(194-198)	2269	0.4878	-	7471B	PPB	08 Jun 2021 17:01:15	U	SMPL
M2625-16 - 1	18-B12(198-202)	1088	0.2535	-	7471B	PPB	08 Jun 2021 17:03:30	U	SMPL
M2625-17 - 1	18-B12(202-206)	218	0.0809	-	7471B	PPB	08 Jun 2021 17:05:45	U	SMPL
M2625-18 - 1	18-B12(206-210)	257	0.0886	-	7471B	PPB	08 Jun 2021 17:08:01	U	SMPL
M2625-19 - 1	18-B12(210-214)	159	0.0692	-	7471B	PPB	08 Jun 2021 17:10:18	U	SMPL
M2625-20 - 1	18-B12(214-218)	89	0.0553	-	7471B	PPB	08 Jun 2021 17:12:35	U	SMPL
PB136939BL - 1	PBS	5	0.0386	-	7471B	PPB	08 Jun 2021 17:14:51	U	SMPL
PB136939BS - 1	LCSS	19053	3.8174	-	7471B	PPB	08 Jun 2021 17:17:07	U	SMPL
CCV54 - 1	CCV54	26763	5.3469	-	7471B	PPB	08 Jun 2021 17:19:22	U	SMPL
CCB54 - 1	CCB54	-75	0.0227	-	7471B	PPB	08 Jun 2021 17:21:37	U	SMPL
M2628-01 - 1	HD-02-060721	3437	0.7195	-	7471B	PPB	08 Jun 2021 17:23:52	U	SMPL
M2628-01DUP - 1	HD-02-060721DUP	3423	0.7167	-	7471B	PPB	08 Jun 2021 17:26:08	U	SMPL
M2628-01MS - 1	HD-02-060721MS	26009	5.1973	-	7471B	PPB	08 Jun 2021 17:28:23	U	SMPL
M2628-01MSD - 1	HD-02-060721MSD	23967	4.7922	-	7471B	PPB	08 Jun 2021 17:30:39	U	SMPL
M2634-01 - 1	R44-STONE-1	22372	4.4758	-	7471B	PPB	08 Jun 2021 17:32:54	U	SMPL
M2634-04 - 1	R44-STONE-2	130579	25.9419	-	7471B	PPB	08 Jun 2021 17:35:10	U	SMPL
M2637-01 - 1	18-B12(230-234)	-313	-0.0245	-	7471B	PPB	08 Jun 2021 17:37:27	U	SMPL
M2637-02 - 1	18-B12(234-238)	81	0.0537	-	7471B	PPB	08 Jun 2021 17:39:50	U	SMPL
M2637-03 - 1	18-B12(238-242)	109	0.0593	-	7471B	PPB	08 Jun 2021 17:42:07	U	SMPL
M2637-04 - 1	18-B12(242-246)	47	0.0470	-	7471B	PPB	08 Jun 2021 17:44:23	U	SMPL
CCV55 - 1	CCV55	23987	4.7962	-	7471B	PPB	09 Jun 2021 08:34:59	U	SMPL
CCB55 - 1	CCB55	-168	0.0043	-	7471B	PPB	09 Jun 2021 08:37:14	U	SMPL
M2637-05 - 1	18-B12(246-248)	291	0.0954	-	7471B	PPB	09 Jun 2021 08:39:29	U	SMPL
M2625-01LX5 - 1		24	0.0424	-	7471B	PPB	09 Jun 2021 08:41:46	U	SMPL
M2625-01A - 1		22072	4.4163	-	7471B	PPB	09 Jun 2021 08:44:02	U	SMPL
M2628-01LX5 - 1		600	0.1567	-	7471B	PPB	09 Jun 2021 08:46:18	U	SMPL
M2628-01A - 1		23194	4.6388	-	7471B	PPB	09 Jun 2021 08:48:35	U	SMPL
M2634-04DLX5 - 1		30978	6.1830	-	7471B	PPB	09 Jun 2021 09:00:06	U	SMPL
CCV56 - 1	CCV56	24665	4.9307	-	7471B	PPB	09 Jun 2021 09:02:22	U	SMPL
CCB56 - 1	CCB56	-59	0.0259	-	7471B	PPB	09 Jun 2021 09:04:40	U	SMPL