



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114853

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	06/08/2021	17:00
2	CAL2	1	Water	NA	NA	21.5	7.00	06/08/2021	17:01
3	CAL3	1	Water	NA	NA	21.4	10.04	06/08/2021	17:02
4	ICV	1	Water	NA	NA	21.3	7.00	06/08/2021	17:05
5	CCV1	1	Water	NA	NA	21.3	2.01	06/08/2021	17:06
6	M2644-06	1	Water	NA	NA	25.8	6.53	06/08/2021	17:10
7	M2644-06DUP	1	Water	NA	NA	25.9	6.54	06/08/2021	17:11
8	M2644-07	1	Water	NA	NA	25.5	7.49	06/08/2021	17:12
9	M2644-09	1	Water	NA	NA	26.2	7.40	06/08/2021	17:15
10	M2645-03	1	Water	NA	NA	26.2	5.17	06/08/2021	17:16
11	M2645-05	1	Water	NA	NA	25.3	5.94	06/08/2021	17:18
12	CCV2	1	Water	NA	NA	21.4	12.02	06/08/2021	17:20

28 114853

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m2644

WorkList ID : 149639

Department : Wet-Chemistry

Date : 06-08-2021 16:43:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/08/2021	Water	M2644-06	pH	Cool 4 deg C	PSEG03	G22	1824	06/08/2021	9040C
06/08/2021	Water	M2644-07	pH	Cool 4 deg C	PSEG03	G22	1825	06/08/2021	9040C
06/08/2021	Water	M2644-09	pH	Cool 4 deg C	PSEG03	G22	1847	06/08/2021	9040C
06/08/2021	Water	M2645-03	pH	Cool 4 deg C	PSEG03	G22	BUR-1041	06/08/2021	9040C
06/08/2021	Water	M2645-05	pH	Cool 4 deg C	PSEG03	G22	BUR-1048	06/08/2021	9040C

Date/Time 06/08/21 16:43
 Raw Sample Received by: 30 wet-chem
 Raw Sample Relinquished by: wet-chem

Date/Time 06/08/21 18:00
 Raw Sample Received by: wet-chem
 Raw Sample Relinquished by: 30 wet-chem