



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114854

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	06/08/2021	17:00
2	CAL2	1	Water	NA	NA	21.5	7.00	06/08/2021	17:01
3	CAL3	1	Water	NA	NA	21.4	10.04	06/08/2021	17:02
4	ICV	1	Water	NA	NA	21.3	7.00	06/08/2021	17:05
5	CCV1	1	Water	NA	NA	21.3	2.01	06/08/2021	17:06
6	M2645-01	1	Solid	20.03	20	26.4	6.69	06/08/2021	17:21
7	M2645-01DUP	1	Solid	20.04	20	26.5	6.70	06/08/2021	17:22
8	CCV2	1	Water	NA	NA	21.4	12.02	06/08/2021	17:23

LB 114854

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s m2645 WorkList ID : 149640 Department : Wet-Chemistry Date : 06-08-2021 16:44:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/08/2021	Solid	M2645-01	pH	Cool 4 deg C	PSEG03	G22	23892	06/08/2021	9045D

Date/Time 06/08/21 1644
Raw Sample Received by: Jo
Raw Sample Relinquished by: AM Wet Chem

Date/Time 06/08/21 18:00
Raw Sample Received by: AM Wet Chem
Raw Sample Relinquished by: JA O'Brien