



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114863

Slope : 99.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2627
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	25.4	4.01	06/09/2021	13:06
2	CAL2	1	Water	NA	NA	25.4	7.01	06/09/2021	13:07
3	CAL3	1	Water	NA	NA	25.5	10.03	06/09/2021	13:09
4	ICV	1	Water	NA	NA	25.6	6.99	06/09/2021	13:11
5	CCV1	1	Water	NA	NA	25.6	2.00	06/09/2021	13:13
6	M2656-01	1	Water	NA	NA	25.4	7.34	06/09/2021	13:15
7	M2665-01	1	Water	NA	NA	25.6	5.17	06/09/2021	13:17
8	M2665-01DUP	1	Water	NA	NA	25.6	5.15	06/09/2021	13:19
9	CCV2	1	Water	NA	NA	25.5	12.02	06/09/2021	13:21

LB114863

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m2656 WorkList ID : 149670 Department : Wet-Chemistry Date : 06-09-2021 13:00:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/09/2021	Water	M2656-01	A pH	Cool 4 deg C	PSEG03	G22	FM628	06/09/2021	9040C
06/03/2021	Water	M2665-01	A pH	Cool 4 deg C	STER15	F51	GILROY-OMEGA-CELL	06/03/2021	9040C

Date/Time 06/09/2021 13:03

Raw Sample Received by: AMCWet-chem

Raw Sample Relinquished by: SD (05-2021)

Date/Time 06/09/2021 13:23

Raw Sample Received by: Jb

Raw Sample Relinquished by: AMCWet-chem