


LB114805

7471B

INSTRUMENT ID: CV1

 μ Abs.:

53225



Linear

A= 0.0000e+000

B= 1.8787e-004

C= -4.2863e-002

Rho= 0.9999161

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% 1)
0.0	0.000	-0.051	-0.051	-45	0.000	-45					
0.2	0.200	0.191	-0.009	1243	0.0 %	1243					
2.5	2.500	2.582	0.082	13974	0.0 %	13974					
5.0	5.000	4.978	-0.022	26727	0.0 %	26727					
7.5	7.500	7.543	0.043	40377	0.0 %	40377					
10.0	10.000	9.957	-0.043	53225	0.0 %	53225					

1
-5
3
0
1
0

LB114865 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Jun 2021 14:12:37

Sample ID	Extended ID	μ Abs.	Conc.	Strnd CondMethod	Units	Date	Type	Type
0.0 - 1	50	-45	-	0.00007471B	PPB	09 Jun 2021 14:28:10	S	Std
0.2 - 1	502	1243	-	0.20007471B	PPB	09 Jun 2021 14:30:30	S	Std
2.5 - 1	52.5	13974	-	2.50007471B	PPB	09 Jun 2021 14:32:47	S	Std
5.0 - 1	55	26727	-	5.00007471B	PPB	09 Jun 2021 14:35:04	S	Std
7.5 - 1	57.5	40377	-	7.50007471B	PPB	09 Jun 2021 14:37:20	S	Std
10.0 - 1	57.0	53225	-	10.00007471B	PPB	09 Jun 2021 14:42:11	S	Std
ICV55 - 1	ICV55	22128	4.1144	-7471B	PPB	09 Jun 2021 14:45:30	U	SMPL
ICB55 - 1	ICB55	-72	-0.0564	-7471B	PPB	09 Jun 2021 14:47:44	U	SMPL
CCV62 - 1	CCV62	26528	4.9411	-7471B	PPB	09 Jun 2021 14:49:59	U	SMPL
CCB62 - 1	CCB62	-68	-0.0556	-7471B	PPB	09 Jun 2021 14:52:13	U	SMPL
CRA - 1	CRA	1210	0.1845	-7471B	PPB	09 Jun 2021 14:54:28	U	SMPL
HighStd - 1	HighStd	53561	10.0199	-7471B	PPB	09 Jun 2021 14:56:43	U	SMPL
ChkStd - 1	ChkStd	37272	6.9596	-7471B	PPB	09 Jun 2021 14:58:58	U	SMPL
PB136956BL - 1	PBS	-119	-0.0652	-7471B	PPB	09 Jun 2021 15:01:15	U	SMPL
M2643-01 - 1	11934	3251	0.5679	-7471B	PPB	09 Jun 2021 15:05:47	U	SMPL
M2643-01DUP - 1	11934DUP	3698	0.6519	-7471B	PPB	09 Jun 2021 15:08:03	U	SMPL
M2643-01MS - 1	11934MS	22208	4.1295	-7471B	PPB	09 Jun 2021 15:12:49	U	SMPL
M2643-01MSD - 1	11934MSD	22295	4.1458	-7471B	PPB	09 Jun 2021 15:15:05	U	SMPL
M2644-04 - 1	329	691	0.0870	-7471B	PPB	09 Jun 2021 15:17:21	U	SMPL
CCV63 - 1	CCV63	26434	4.9234	-7471B	PPB	09 Jun 2021 15:19:36	U	SMPL
CCB63 - 1	CCB63	-79	-0.0577	-7471B	PPB	09 Jun 2021 15:21:51	U	SMPL
M2645-01 - 1	23892	24	-0.0384	-7471B	PPB	09 Jun 2021 15:24:07	U	SMPL
M2646-01 - 1	STONE-COMP-1	492	0.0496	-7471B	PPB	09 Jun 2021 15:26:22	U	SMPL
M2646-03 - 1	STONE-COMP-2	382	0.0289	-7471B	PPB	09 Jun 2021 15:28:38	U	SMPL
M2647-01 - 1	SU-01-060821	1366	0.2138	-7471B	PPB	09 Jun 2021 15:30:53	U	SMPL
M2648-01 - 1	CRUSH-R	621	0.0738	-7471B	PPB	09 Jun 2021 15:33:09	U	SMPL
M2650-01 - 1	DELAWANNA-RT10-STOCKPILE-1	3860	0.6823	-7471B	PPB	09 Jun 2021 15:35:25	U	SMPL
M2650-02 - 1	DELAWANNA-RT10-STOCKPILE-2	3861	0.6825	-7471B	PPB	09 Jun 2021 15:37:41	U	SMPL
M2650-03 - 1	DELAWANNA-RT10-STOCKPILE-3	307	0.0148	-7471B	PPB	09 Jun 2021 15:39:58	U	SMPL
M2651-01 - 1	CON-ED-EXCAVATION	10422	1.9152	-7471B	PPB	09 Jun 2021 15:42:15	U	SMPL
M2651-02 - 1	17-B2(0-5)	54701	10.2341	-7471B	PPB	09 Jun 2021 15:44:30	U	SMPL
CCV64 - 1	CCV64	26409	4.9187	-7471B	PPB	09 Jun 2021 15:46:46	U	SMPL
CCB64 - 1	CCB64	-95	-0.0607	-7471B	PPB	09 Jun 2021 15:49:02	U	SMPL
M2651-03 - 1	DEP-CRANE-PAD-EXCAVATION	10419	1.9146	-7471B	PPB	09 Jun 2021 15:51:18	U	SMPL
M2651-04 - 1	17-B6(0-4)	9858	1.8092	-7471B	PPB	09 Jun 2021 15:53:33	U	SMPL
M2651-05 - 1	17-B6(4-8)	4825	0.8636	-7471B	PPB	09 Jun 2021 15:55:49	U	SMPL
PB136956BS - 1	LCSS	22337	4.1537	-7471B	PPB	09 Jun 2021 15:58:05	U	SMPL
M2643-01LX5 - 1		552	0.0608	-7471B	PPB	09 Jun 2021 16:00:21	U	SMPL
M2643-01A - 1		24872	4.6300	-7471B	PPB	09 Jun 2021 16:02:36	U	SMPL
M2651-02DLX2 - 1		26543	4.9439	-7471B	PPB	09 Jun 2021 16:04:53	U	SMPL
CCV65 - 1	CCV65	26635	4.9612	-7471B	PPB	09 Jun 2021 16:07:10	U	SMPL
CCB65 - 1	CCB65	-90	-0.0598	-7471B	PPB	09 Jun 2021 16:09:26	U	SMPL