



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114867

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	06/09/2021	14:25
2	CAL2	1	Water	NA	NA	21.3	7.00	06/09/2021	14:26
3	CAL3	1	Water	NA	NA	21.3	10.03	06/09/2021	14:27
4	ICV	1	Water	NA	NA	21.4	7.02	06/09/2021	14:29
5	CCV1	1	Water	NA	NA	21.4	2.01	06/09/2021	14:30
6	M2658-01	1	Solid	20.01	20	25.9	11.46	06/09/2021	14:35
7	M2658-04	1	Solid	20.02	20	25.6	10.48	06/09/2021	14:37
8	M2658-07	1	Solid	20.03	20	25.4	11.60	06/09/2021	14:40
9	M2658-10	1	Solid	20.04	20	25.1	11.35	06/09/2021	14:42
10	M2658-13	1	Solid	20.02	20	24.7	10.64	06/09/2021	14:45
11	M2658-16	1	Solid	20.04	20	24.3	10.19	06/09/2021	14:46
12	M2658-16DUP	1	Solid	20.03	20	24.4	10.20	06/09/2021	14:47
13	CCV2	1	Water	NA	NA	21.6	12.02	06/09/2021	14:48

WORKLIST(Hardcopy Internal Chain)

1B 114881
1B 114861

WorkList Name : ph m2638

WorkList ID : 149678

Department : Wet-Chemistry

Date : 06-09-2021 14:11:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/09/2021	Solid	M2658-01	pH	Cool 4 deg C	PSEG03	G31	T-FLOOR	06/09/2021	9045D
06/09/2021	Solid	M2658-04	pH	Cool 4 deg C	PSEG03	G31	P-WALL	06/09/2021	9045D
06/09/2021	Solid	M2658-07	pH	Cool 4 deg C	PSEG03	G31	G-FLOOR-1	06/09/2021	9045D
06/09/2021	Solid	M2658-10	pH	Cool 4 deg C	PSEG03	G31	G-FLOOR-2	06/09/2021	9045D
06/09/2021	Solid	M2658-13	pH	Cool 4 deg C	PSEG03	G31	G-WALL-1	06/09/2021	9045D
06/09/2021	Solid	M2658-16	pH	Cool 4 deg C	PSEG03	G31	G-WALL-2	06/09/2021	9045D

Date/Time 06/09/21 14:20
Raw Sample Received by: TO GUY (54)
Raw Sample Relinquished by: TO GUY (54)

Date/Time 06/09/21 17:30
Raw Sample Received by: TO GUY (54)
Raw Sample Relinquished by: TO GUY (54)