


LB114873

SFAM01.1

INSTRUMENT ID: CV2

 μ Abs.:

47455



Linear

A= 0.0000e+000

B= 2.1077e-004

C= -6.3791e-002

Rho= 0.9999410

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.019	-0.019	212	0.000	212					=
0.05	0.050					204					-3
0.20	0.200	0.194	-0.006	1222	0.0 %	1222					-1
2.50	2.500	2.486	-0.014	12096	0.0 %	12096					1
5.00	5.000	5.046	0.046	24245	0.0 %	24245					1
7.50	7.500	7.555	0.055	36150	0.0 %	36150					1
10.0	10.000	9.938	-0.062	47455	0.0 %	47455					-1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	212	-		0 SFAM01.1	PPB	6/9/2021 16:22	Std
	0.2 S01	1222	-		0.2 SFAM01.1	PPB	6/9/2021 16:24	Std
	2.5 S02	12096	-		2.5 SFAM01.1	PPB	6/9/2021 16:27	Std
	5 S03	24245	-		5 SFAM01.1	PPB	6/9/2021 16:29	Std
	7.5 S04	36150	-		7.5 SFAM01.1	PPB	6/9/2021 16:31	Std
	10 S05	47455	-		10 SFAM01.1	PPB	6/9/2021 16:33	Std
ICV038	ICV038	18646	3.8661 -		SFAM01.1	PPB	6/9/2021 16:36	SMPL
ICB038	ICB038	143	-0.0337 -		SFAM01.1	PPB	6/9/2021 16:38	SMPL
CCV066	CCV066	24278	5.0532 -		SFAM01.1	PPB	6/9/2021 16:41	SMPL
CCB066	CCB066	132	-0.036 -		SFAM01.1	PPB	6/9/2021 16:43	SMPL
PB136970BL	PBS970	307	0.0009 -		SFAM01.1	PPB	6/9/2021 16:45	SMPL
M2636-19	ESQK0	1643	0.2825 -		SFAM01.1	PPB	6/9/2021 16:52	SMPL
M2636-20	ESQK0D	1603	0.2741 -		SFAM01.1	PPB	6/9/2021 16:55	SMPL
M2636-21	ESQK0S	13310	2.7415 -		SFAM01.1	PPB	6/9/2021 16:57	SMPL
M2636-01	ESQH2	1179	0.1847 -		SFAM01.1	PPB	6/9/2021 16:59	SMPL
M2636-02	ESQH3	1186	0.1862 -		SFAM01.1	PPB	6/9/2021 17:02	SMPL
M2636-03	ESQH4	1754	0.3059 -		SFAM01.1	PPB	6/9/2021 17:04	SMPL
M2636-04	ESQH5	2035	0.3651 -		SFAM01.1	PPB	6/9/2021 17:06	SMPL
M2636-05	ESQH6	1469	0.2458 -		SFAM01.1	PPB	6/9/2021 17:08	SMPL
M2636-06	ESQH7	1504	0.2532 -		SFAM01.1	PPB	6/9/2021 17:11	SMPL
M2636-07	ESQH8	1991	0.3558 -		SFAM01.1	PPB	6/9/2021 17:13	SMPL
M2636-08	ESQH9	1352	0.2212 -		SFAM01.1	PPB	6/9/2021 17:15	SMPL
M2636-09	ESQJ0	1677	0.2897 -		SFAM01.1	PPB	6/9/2021 17:17	SMPL
M2636-10	ESQJ1	1203	0.1898 -		SFAM01.1	PPB	6/9/2021 17:20	SMPL
M2636-11	ESQJ2	1403	0.2319 -		SFAM01.1	PPB	6/9/2021 17:22	SMPL
M2636-12	ESQJ3	1849	0.3259 -		SFAM01.1	PPB	6/9/2021 17:24	SMPL
M2636-13	ESQJ4	1182	0.1853 -		SFAM01.1	PPB	6/9/2021 17:26	SMPL
M2636-14	ESQJ5	1346	0.2199 -		SFAM01.1	PPB	6/9/2021 17:29	SMPL
M2636-15	ESQJ6	1568	0.2667 -		SFAM01.1	PPB	6/9/2021 17:31	SMPL
M2636-16	ESQJ7	1612	0.276 -		SFAM01.1	PPB	6/9/2021 17:33	SMPL
CCV067	CCV067	24447	5.0888 -		SFAM01.1	PPB	6/9/2021 17:35	SMPL
CCB067	CCB067	136	-0.0351 -		SFAM01.1	PPB	6/9/2021 17:38	SMPL
M2636-17	ESQJ8	1821	0.32 -		SFAM01.1	PPB	6/9/2021 17:40	SMPL
M2636-18	ESQJ9	1616	0.2768 -		SFAM01.1	PPB	6/9/2021 17:42	SMPL
CCV068	CCV068	24403	5.0795 -		SFAM01.1	PPB	6/9/2021 17:44	SMPL
CCB068	CCB068	138	-0.0347 -		SFAM01.1	PPB	6/9/2021 17:47	SMPL