



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114874

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/09/2021	15:10
2	M2645-03		1	8	100.00	>212.0	>212.0	06/09/2021	15:35
3	M2645-03DUP		1	8	100.00	>212.0	>212.0	06/09/2021	16:05
4	M2645-05		1	9	100.00	>212.0	>212.0	06/09/2021	16:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 114874

WorkList Name : fp6921 WorkList ID : 149662 Department : Wet-Chemistry Date : 06-09-2021 09:07:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/21/2021	Water	M2645-03	Flash Point	Cool 4 deg C	PSEG03	G22	BUR-1041	06/08/2021	1010A
06/21/2021	Water	M2645-05	Flash Point	Cool 4 deg C	PSEG03	G22	BUR-1048	06/08/2021	1010A

Date/Time 06/09/21 14:45
 Raw Sample Received by: KS Guechuan
 Raw Sample Relinquished by: CP 3M

Date/Time 06/09/21 16:35
 Raw Sample Received by: CP 3M
 Raw Sample Relinquished by: KS Guechuan