

Analytical Summary ReportAnalysis Method: 1010AReviewed By: ketankumarParameter: Flash PointSupervisor Review By: apatelRun Number: LB114882Ambient Barometric Pressure (mmHg): 760.00Thermometer ID: flashpointBarometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	06/10/2021	09:35
2	M2659-01		1	9	95.00	203.0	203.0	06/10/2021	10:05
3	M2659-01DUP		1	9	96.00	204.8	204.8	06/10/2021	10:40

Result = (Celsius * 1.8) + 32**Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06**

WORKLIST(Hardcopy Internal Chain)

LB114882

WorkList Name : fp61021-1 WorkList ID : 149704 Department : Wet-Chemistry Date : 06-10-2021 09:02:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/22/2021	Solid	M2659-01	Flash Point	Cool 4 deg C	PSEG03	G31	31454	06/09/2021	1010A

Date/Time 06/10/21 09:05
 Raw Sample Received by: KS (Wehchey)
 Raw Sample Relinquished by: SA GY (non)

Date/Time 06/10/21 10:45
 Raw Sample Received by: SA GY (non)
 Raw Sample Relinquished by: KS (Wehchey)