



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114883

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID:

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	06/10/2021	09:35
2	M2656-01		1	9	100.00	>212.0	>212.0	06/10/2021	11:10
3	M2656-01DUP		1	9	100.00	>212.0	>212.0	06/10/2021	11:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114883

WorkList Name : fp61021 WorkList ID : 149703 Department : Wet-Chemistry Date : 06-10-2021 09:01:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/14/2021	Water	M2656-01	Flash Point	Cool 4 deg C	PSEG03	G22	FM628	06/09/2021	1010A

Date/Time 06/10/21 09:05
 Raw Sample Received by: RS (Wetchem)
 Raw Sample Relinquished by: JS (Wetchem)

Date/Time 06/10/21 11:40
 Raw Sample Received by: JS (Wetchem)
 Raw Sample Relinquished by: JS (Wetchem)