



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114893

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	06/10/2021	12:20
2	CAL2	1	Water	NA	NA	21.1	7.01	06/10/2021	12:21
3	CAL3	1	Water	NA	NA	21.1	10.02	06/10/2021	12:22
4	ICV	1	Water	NA	NA	21.2	6.99	06/10/2021	12:23
5	CCV1	1	Water	NA	NA	21.1	2.01	06/10/2021	12:25
6	M2666-01	1	Solid	20.03	20	24.4	8.62	06/10/2021	12:26
7	M2666-01DUP	1	Solid	20.04	20	24.3	8.63	06/10/2021	12:27
8	M2674-01	1	Solid	20.02	20	24.3	5.40	06/10/2021	12:30
9	M2674-02	1	Solid	20.03	20	24.2	5.36	06/10/2021	12:31
10	M2674-03	1	Solid	20.04	20	24.2	5.02	06/10/2021	12:33
11	M2674-04	1	Solid	20.02	20	24.1	5.09	06/10/2021	12:35
12	M2674-05	1	Solid	20.03	20	23.7	5.18	06/10/2021	12:36
13	M2674-06	1	Solid	20.04	20	23.8	5.22	06/10/2021	12:37
14	CCV2	1	Water	NA	NA	21.2	12.02	06/10/2021	12:39

✓ 114893

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH M2674 WorkList ID: 149714 Department: Wet-Chemistry Date: 06-10-2021 12:12:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/09/2021	Solid	M2666-01	pH	Cool 4 deg C	NEWP02	G31	RIVER-ST-TO-CAMDEN-	06/09/2021	9045D
06/09/2021	Solid	M2674-01	pH	Cool 4 deg C	ENVI77	F62	TMR-DS-01-060921	06/09/2021	9045D
06/09/2021	Solid	M2674-02	pH	Cool 4 deg C	ENVI77	F62	TMR-DS-02-060921	06/09/2021	9045D
06/09/2021	Solid	M2674-03	pH	Cool 4 deg C	ENVI77	F62	TMR-DS-03-060921	06/09/2021	9045D
06/09/2021	Solid	M2674-04	pH	Cool 4 deg C	ENVI77	F62	TMR-DS-04-060921	06/09/2021	9045D
06/09/2021	Solid	M2674-05	pH	Cool 4 deg C	ENVI77	F62	TMR-DS-05-060921	06/09/2021	9045D
06/09/2021	Solid	M2674-06	pH	Cool 4 deg C	ENVI77	F62	TMR-DS-06-060921	06/09/2021	9045D

Date/Time 06-10-21 12:15
 Raw Sample Received by: PHCUT-060921
 Raw Sample Relinquished by: CL SM

Date/Time 06-10-21 14:00
 Raw Sample Received by: CL SM
 Raw Sample Relinquished by: JACUT-060921