


LB114901

245.1

INSTRUMENT ID: CV2



A= 0.0000e+000

B= 2.5857e-004

C= -7.7125e-003

Rho= 0.9998110

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.0	0.000	-0.002	-0.002	21	0.000	21					-
0.05	0.050	0.051	0.001	226	0.0 %	226	0				2
0.2	0.200	0.206	0.006	827	0.0 %	827					3
2.5	2.500	2.459	-0.041	9541	0.0 %	9541					-2
5.0	5.000	5.125	0.125	19851	0.0 %	19851					3
7.5	7.500	7.369	-0.131	28527	0.0 %	28527					-2
10.0	10.000	10.042	0.042	38868	0.0 %	38868					0

LB114901 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Jun 2021 10:44:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.0000	245.1	PPB	11 Jun 2021 10:54:19	S	Std
0.05 - 1	50.05	226	-	0.0500	245.1	PPB	11 Jun 2021 10:56:36	S	Std
0.2 - 1	50.2	827	-	0.2000	245.1	PPB	11 Jun 2021 10:58:53	S	Std
2.5 - 1	52.5	9541	-	2.5000	245.1	PPB	11 Jun 2021 11:01:09	S	Std
5.0 - 1	55	19851	-	5.0000	245.1	PPB	11 Jun 2021 11:03:25	S	Std
7.5 - 1	57.5	28527	-	7.5000	245.1	PPB	11 Jun 2021 11:05:41	S	Std
10.0 - 1	510	38868	-	10.0000	245.1	PPB	11 Jun 2021 11:28:18	S	Std
ICV58 - 1	ICV58	15146	3.9086	-245.1		PPB	11 Jun 2021 11:31:07	U	SMPL
ICB58 - 1	ICB58	-45	-0.0193	-245.1		PPB	11 Jun 2021 11:33:21	U	SMPL
CCV76 - 1	CCV76	19820	5.1172	-245.1		PPB	11 Jun 2021 11:35:36	U	SMPL
CCB76 - 1	CCB76	-55	-0.0219	-245.1		PPB	11 Jun 2021 11:37:51	U	SMPL
CRA - 1	CRA	198	0.0435	-245.1		PPB	11 Jun 2021 11:40:06	U	SMPL
HighStd - 1	HighStd	38785	10.0210	-245.1		PPB	11 Jun 2021 11:42:21	U	SMPL
ChkStd - 1	ChkStd	26683	6.8917	-245.1		PPB	11 Jun 2021 11:44:36	U	SMPL
PB137018BL - 1	PBW	-105	-0.0349	-245.1		PPB	11 Jun 2021 11:46:53	U	SMPL
PB137018BS - 1	LCSW	15821	4.0831	-245.1		PPB	11 Jun 2021 11:49:09	U	SMPL
M2262-10 - 1	LOD-MDL-WATER-04-QT2-2021	190	0.0414	-245.1		PPB	11 Jun 2021 11:54:01	U	SMPL
M2262-11 - 1	LOQ-WATER-05-QT2-2021	784	0.1950	-245.1		PPB	11 Jun 2021 11:56:18	U	SMPL
M2586-02 - 1	COMP	139	0.0282	-245.1		PPB	11 Jun 2021 11:58:34	U	SMPL
M2681-02 - 1	EFFLUENT-METALS	303	0.0706	-245.1		PPB	11 Jun 2021 12:00:52	U	SMPL
CCV77 - 1	CCV77	19705	5.0874	-245.1		PPB	11 Jun 2021 12:03:07	U	SMPL
CCB77 - 1	CCB77	-68	-0.0253	-245.1		PPB	11 Jun 2021 12:05:22	U	SMPL
M2681-04 - 1	EFFLUENT-METALS	262	0.0600	-245.1		PPB	11 Jun 2021 12:07:37	U	SMPL
M2681-04DUP - 1	EFFLUENT-METALS DUP	217	0.0484	-245.1		PPB	11 Jun 2021 12:09:53	U	SMPL
M2681-04MS - 1	EFFLUENT-METALS MS	16672	4.3032	-245.1		PPB	11 Jun 2021 12:12:08	U	SMPL
M2681-04MSD - 1	EFFLUENT-METALS MSD	15990	4.1268	-245.1		PPB	11 Jun 2021 12:14:24	U	SMPL
M2681-04LX5 - 1		11	-0.0049	-245.1		PPB	11 Jun 2021 12:16:39	U	SMPL
M2681-04A - 1		16194	4.1796	-245.1		PPB	11 Jun 2021 12:18:55	U	SMPL
CCV78 - 1	CCV78	18423	4.7559	-245.1		PPB	11 Jun 2021 12:21:11	U	SMPL
CCB78 - 1	CCB78	-157	-0.0483	-245.1		PPB	11 Jun 2021 12:23:27	U	SMPL