

LB114922


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4868E



A= 0.0000e+000

B= 2.0354e-004

C= -1.9624e-002

Rho= 0.9998091

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Y/D
0.0	0.000	-0.017	-0.017	15	0.000	15					-
0.2	0.200	0.191	-0.009	1037	0.0 %	1037					-5
2.5	2.500	2.460	-0.040	12182	0.0 %	12182					-2
5.0	5.000	5.057	0.057	24943	0.0 %	24943					1
7.5	7.500	7.618	0.118	37522	0.0 %	37522					2
10.0	10.000	9.890	-0.110	48686	0.0 %	48686					-1

LB114922 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Jun 2021 16:28:46

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	15	-	0.00007471B	PPB		11 Jun 2021 16:32:57	S	Std
0.2 - 1	SD.2	1037	-	0.20007471B	PPB		11 Jun 2021 16:35:13	S	Std
2.5 - 1	J2-T	12182	-	2.50007471B	PPB		11 Jun 2021 16:37:30	S	Std
5.0 - 1	S5	24943	-	5.00007471B	PPB		11 Jun 2021 16:39:47	S	Std
7.5 - 1	S7.5	37522	-	7.50007471B	PPB		11 Jun 2021 16:44:38	S	Std
10.0 - 1	S10	48686	-	10.00007471B	PPB		11 Jun 2021 16:46:54	S	Std
ICV59 - 1	ICV59	19211	3.8906	-7471B	PPB		11 Jun 2021 16:49:42	U	SMPL
ICB59 - 1	ICB59	-6	-0.0208	-7471B	PPB		11 Jun 2021 16:55:13	U	SMPL
CCV83 - 1	CCV83	24469	4.9609	-7471B	PPB		11 Jun 2021 16:57:27	U	SMPL
CCB83 - 1	CCB83	-20	-0.0237	-7471B	PPB		11 Jun 2021 17:02:23	U	SMPL
CRA - 1	CRA	986	0.1811	-7471B	PPB		11 Jun 2021 17:04:37	U	SMPL
HighStd - 1	HighStd	48008	9.7521	-7471B	PPB		11 Jun 2021 17:06:52	U	SMPL
ChkStd - 1	ChkStd	33953	6.8913	-7471B	PPB		11 Jun 2021 17:09:07	U	SMPL
PB136982BL - 1	PBS	-52	-0.0302	-7471B	PPB		11 Jun 2021 17:16:11	U	SMPL
PB136982BS - 1	LCSS	18868	3.8208	-7471B	PPB		11 Jun 2021 17:18:27	U	SMPL
M2638-02 - 1	ND-1	1953	0.3779	-7471B	PPB		11 Jun 2021 17:20:43	U	SMPL
M2638-02DUP - 1	ND-1DUP	1826	0.3520	-7471B	PPB		11 Jun 2021 17:23:01	U	SMPL
M2638-02MS - 1	ND-1MS	19726	3.9955	-7471B	PPB		11 Jun 2021 17:25:17	U	SMPL
M2638-02MSD - 1	ND-1MSD	21741	4.4056	-7471B	PPB		11 Jun 2021 17:27:34	U	SMPL
M2660-01 - 1	OR-03-060921	980	0.1798	-7471B	PPB		11 Jun 2021 17:29:49	U	SMPL
CCV84 - 1	CCV84	23639	4.7919	-7471B	PPB		11 Jun 2021 17:32:06	U	SMPL
CCB84 - 1	CCB84	3	-0.0190	-7471B	PPB		11 Jun 2021 17:37:01	U	SMPL
M2262-01 - 1	LOD-MDL-SOIL-01-QT2-2021	850	0.1534	-7471B	PPB		11 Jun 2021 17:39:16	U	SMPL
M2262-03 - 1	MDL-SOIL-03-QT2-2021	264	0.0341	-7471B	PPB		11 Jun 2021 17:41:31	U	SMPL
M2661-01 - 1	OK-01-060921	1777	0.3421	-7471B	PPB		11 Jun 2021 17:46:16	U	SMPL
M2666-01 - 1	RIVER-ST-TO-CAMDEN-COMP	7320	1.4703	-7471B	PPB		11 Jun 2021 17:48:31	U	SMPL
M2673-01 - 1	17-B6(88-92)	226	0.0264	-7471B	PPB		11 Jun 2021 17:50:47	U	SMPL
M2673-02 - 1	17-B6(92-96)	650	0.1127	-7471B	PPB		11 Jun 2021 17:53:03	U	SMPL
M2673-03 - 1	17-B6(96-100)	499	0.0819	-7471B	PPB		11 Jun 2021 17:55:20	U	SMPL
M2674-01 - 1	TMR-DS-01-060921	2622	0.5141	-7471B	PPB		11 Jun 2021 17:57:37	U	SMPL
M2674-02 - 1	TMR-DS-02-060921	2821	0.5546	-7471B	PPB		11 Jun 2021 17:59:53	U	SMPL
M2674-03 - 1	TMR-DS-03-060921	2468	0.4827	-7471B	PPB		11 Jun 2021 18:02:09	U	SMPL
CCV85 - 1	CCV85	23667	4.7976	-7471B	PPB		11 Jun 2021 18:04:24	U	SMPL
CCB85 - 1	CCB85	-105	-0.0410	-7471B	PPB		11 Jun 2021 18:06:39	U	SMPL
M2674-04 - 1	TMR-DS-04-060921	2371	0.4630	-7471B	PPB		11 Jun 2021 18:08:55	U	SMPL
M2674-05 - 1	TMR-DS-05-060921	1906	0.3683	-7471B	PPB		11 Jun 2021 18:11:11	U	SMPL
M2674-06 - 1	TMR-DS-06-060921	2719	0.5338	-7471B	PPB		11 Jun 2021 18:13:26	U	SMPL
M2638-02LX5 - 1	MDL-SOIL-03-QT2-2021	305	0.0425	-7471B	PPB		11 Jun 2021 18:15:42	U	SMPL
M2638-02A - 1		21543	4.3653	-7471B	PPB		11 Jun 2021 18:17:58	U	SMPL
CCV86 - 1	CCV86	23584	4.7807	-7471B	PPB		11 Jun 2021 18:20:14	U	SMPL
CCB86 - 1	CCB86	98	0.0003	-7471B	PPB		11 Jun 2021 18:24:48	U	SMPL