

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB114932

ANALYST: ketankumar
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP90938
IODINE SOLUTION .025N 1L	W2763
Starch Solution, 4L	W2703

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB137051BL		1	5.00	50	2.00	0.00	1.90	1.90	0.10	0.00	0.00	06/12/2021	13:40
2	M2639-02		1	5.05	50	2.00	1.90	3.76	1.86	0.14	0.04	3.17	06/12/2021	13:42
3	M2639-02DUP		1	5.04	50	2.00	3.76	5.64	1.88	0.12	0.02	1.59	06/12/2021	13:45
4	M2687-11		1	5.05	50	2.00	5.64	7.48	1.84	0.16	0.06	4.75	06/12/2021	13:48
5	M2687-12		1	5.04	50	2.00	0.00	1.82	1.82	0.18	0.08	6.35	06/12/2021	13:50
6	M2687-13		1	5.02	50	2.00	1.82	3.68	1.86	0.14	0.04	3.19	06/12/2021	13:52
7	M2687-14		1	5.03	50	2.00	3.68	5.52	1.84	0.16	0.06	4.77	06/12/2021	13:55
8	M2687-15		1	5.03	50	2.00	5.52	7.40	1.88	0.12	0.02	1.59	06/12/2021	13:58
9	M2687-16		1	5.04	50	2.00	0.00	1.86	1.86	0.14	0.04	3.17	06/12/2021	14:00
10	M2687-17		1	5.02	50	2.00	1.86	3.70	1.84	0.16	0.06	4.78	06/12/2021	14:02
11	M2687-18		1	5.04	50	2.00	3.70	5.50	1.80	0.20	0.10	7.94	06/12/2021	14:05
12	M2687-19		1	5.05	50	2.00	5.50	7.32	1.82	0.18	0.08	6.34	06/12/2021	14:08

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB114932

ANALYST: ketankumar
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP90938
IODINE SOLUTION .025N 1L	W2763
Starch Solution, 4L	W2703

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
13	M2687-20		1	5.03	50	2.00	0.00	1.84	1.84	0.16	0.06	4.77	06/12/2021	14:10

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume