


LB114940

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.5399e-004 *slope*C= 4.7493e-003 *y-intercept*

Rho= 0.9999174

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>C/D</i>
0.00	0.000	0.009	0.009	15	0.000	15					<i>2</i>
0.05	0.050	0.051	0.001	183	0.0 %	183					<i>4</i>
0.20	0.200	0.208	0.008	801	0.0 %	801					<i>-2</i>
2.50	2.500	2.460	-0.040	9667	0.0 %	9667					<i>1</i>
5.00	5.000	5.071	0.071	19948	0.0 %	19948					<i>-1</i>
7.50	7.500	7.413	-0.087	29166	0.0 %	29166					<i>0</i>
10.0	10.000	10.038	0.038	39504	0.0 %	39504					

LB114940
INSTRUMENT ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		15 -		0 SFAM01.1	PPB	6/13/2021 15:04	Std
0.05 S01		183 -		0.05 SFAM01.1	PPB	6/13/2021 15:06	Std
0.2 S02		801 -		0.2 SFAM01.1	PPB	6/13/2021 15:08	Std
2.5 S03		9667 -		2.5 SFAM01.1	PPB	6/13/2021 15:10	Std
5 S04		19948 -		5 SFAM01.1	PPB	6/13/2021 15:13	Std
7.5 S05		29166 -		7.5 SFAM01.1	PPB	6/13/2021 15:15	Std
10 S06		39504 -		10 SFAM01.1	PPB	6/13/2021 15:17	Std
ICV041	ICV041	15788	4.0147 -	SFAM01.1	PPB	6/13/2021 15:20	SMPL
ICB041	ICB041	-18	0.0002 -	SFAM01.1	PPB	6/13/2021 15:25	SMPL
CCV092	CCV092	20160	5.1251 -	SFAM01.1	PPB	6/13/2021 15:27	SMPL
CCB092	CCB092	-14	0.0012 -	SFAM01.1	PPB	6/13/2021 15:32	SMPL
PB137071E	PBW071	-6	0.0032 -	SFAM01.1	PPB	6/13/2021 15:34	SMPL
M2250-01	MDL-WATE	179	0.0502 -	SFAM01.1	PPB	6/13/2021 15:36	SMPL
BLANK-02	BLANK-02	9	0.007 -	SFAM01.1	PPB	6/13/2021 15:39	SMPL
CCV093	CCV093	19794	5.0321 -	SFAM01.1	PPB	6/13/2021 15:41	SMPL
CCB093	CCB093	-7	0.003 -	SFAM01.1	PPB	6/13/2021 15:46	SMPL