


LB114941

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.2794e-004 *slope*C= 1.5658e-002 *y-intercept*

Rho= 0.9999749

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | <i>%ob</i> |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|------------|
| 0.00   | 0.000  | 0.020  | 0.020  | 17    | 0.000      | 17    |       |       |       |       | -          |
| 0.05   | 0.050  | 0.062  | 0.012  | 203   | 0.0 %      | 203   |       |       |       |       | 24         |
| 0.20   | 0.200  | 0.214  | 0.014  | 868   | 0.0 %      | 868   |       |       |       |       | 7          |
| 2.50   | 2.500  | 2.441  | -0.059 | 10641 | 0.0 %      | 10641 |       |       |       |       | -2         |
| 5.00   | 5.000  | 4.988  | -0.012 | 21816 | 0.0 %      | 21816 |       |       |       |       | 0          |
| 7.50   | 7.500  | 7.523  | 0.023  | 32936 | 0.0 %      | 32936 |       |       |       |       | 0          |
| 10.0   | 10.000 | 10.002 | 0.002  | 43813 | 0.0 %      | 43813 |       |       |       |       | 0          |

LB114941  
INSTRUMENT ID : CV1

| Sample ID | Extended II $\mu$ Abs. | Conc.   | Std Conc  | Method        | Units | Date            | Type |
|-----------|------------------------|---------|-----------|---------------|-------|-----------------|------|
| 0         |                        | 17 -    |           | 0 SFAM01.1    | PPB   | 6/13/2021 16:34 | Std  |
| 0.05      |                        | 203 -   |           | 0.05 SFAM01.1 | PPB   | 6/13/2021 16:37 | Std  |
| 0.2       |                        | 868 -   |           | 0.2 SFAM01.1  | PPB   | 6/13/2021 16:39 | Std  |
| 2.5       |                        | 10641 - |           | 2.5 SFAM01.1  | PPB   | 6/13/2021 16:41 | Std  |
| 5         |                        | 21816 - |           | 5 SFAM01.1    | PPB   | 6/13/2021 16:44 | Std  |
| 7.5       |                        | 32936 - |           | 7.5 SFAM01.1  | PPB   | 6/13/2021 16:46 | Std  |
| 10        |                        | 43813 - |           | 10 SFAM01.1   | PPB   | 6/13/2021 16:53 | Std  |
| ICV042    | ICV042                 | 17284   | 3.9554 -  | SFAM01.1      | PPB   | 6/13/2021 16:56 | SMPL |
| ICB042    | ICB042                 | -147    | -0.0178 - | SFAM01.1      | PPB   | 6/13/2021 16:58 | SMPL |
| CCV094    | CCV094                 | 22449   | 5.1327 -  | SFAM01.1      | PPB   | 6/13/2021 17:01 | SMPL |
| CCB094    | CCB094                 | -156    | -0.0199 - | SFAM01.1      | PPB   | 6/13/2021 17:03 | SMPL |
| PB137075E | PBS075                 | -11     | 0.0132 -  | SFAM01.1      | PPB   | 6/13/2021 17:05 | SMPL |
| M2250-03  | MDL-SOIL-0             | 192     | 0.0594 -  | SFAM01.1      | PPB   | 6/13/2021 17:07 | SMPL |
| BLANK-04  | BLANK-04               | -4      | 0.0147 -  | SFAM01.1      | PPB   | 6/13/2021 17:10 | SMPL |
| CCV095    | CCV095                 | 22385   | 5.1181 -  | SFAM01.1      | PPB   | 6/13/2021 17:12 | SMPL |
| CCB095    | CCB095                 | -122    | -0.0122 - | SFAM01.1      | PPB   | 6/13/2021 17:14 | SMPL |