


LB114942

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.9067e-004

C= 5.7568e-002

Rho= 0.9999260

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.061	0.061	20	0.000	20					
0.2	0.200	0.191	-0.009	700	0.0 %	700					
2.5	2.500	2.436	-0.064	12476	0.0 %	12476					
5.0	5.000	4.966	-0.034	25741	0.0 %	25741					
7.5	7.500	7.554	0.054	39317	0.0 %	39317					
10.0	10.000	9.991	-0.009	52100	0.0 %	52100					

4/00
 -5
 -3
 -1
 1
 0

LB114942 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Jun 2021 10:59:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	20	-	0.00007471B	PPB		14 Jun 2021 11:02:12	S	Std
0.2 - 1	50.2	700	-	0.20007471B	PPB		14 Jun 2021 11:04:28	S	Std
2.5 - 1	52.5	12476	-	2.50007471B	PPB		14 Jun 2021 11:06:45	S	Std
5.0 - 1	55	25741	-	5.00007471B	PPB		14 Jun 2021 11:09:02	S	Std
7.5 - 1	57.5	39317	-	7.50007471B	PPB		14 Jun 2021 11:11:18	S	Std
10.0 - 1	57.5	52100	-	10.00007471B	PPB		14 Jun 2021 11:13:35	S	Std
ICV61 - 1	ICV61	20403	3.9478	-7471B	PPB		14 Jun 2021 11:16:40	U	SMPL
ICB61 - 1	ICB61	-139	0.0311	-7471B	PPB		14 Jun 2021 11:18:55	U	SMPL
CCV87 - 1	CCV87	26198	5.0527	-7471B	PPB		14 Jun 2021 11:21:11	U	SMPL
CCB87 - 1	CCB87	-97	0.0391	-7471B	PPB		14 Jun 2021 11:23:25	U	SMPL
CRA - 1	CRA	655	0.1825	-7471B	PPB		14 Jun 2021 11:25:42	U	SMPL
HighStd - 1	HighStd	50756	9.7352	-7471B	PPB		14 Jun 2021 11:27:57	U	SMPL
ChkStd - 1	ChkStd	36467	7.0107	-7471B	PPB		14 Jun 2021 11:30:12	U	SMPL
PB137079BL - 1	PBS	-235	0.0128	-7471B	PPB		14 Jun 2021 11:32:29	U	SMPL
PB137079BS - 1	LCSS	19825	3.8376	-7471B	PPB		14 Jun 2021 11:34:47	U	SMPL
M2672-01 - 1	17-B6(8-12)	3286	0.6841	-7471B	PPB		14 Jun 2021 11:37:03	U	SMPL
M2672-01DUP - 1	17-B6(8-12)DUP	2687	0.5699	-7471B	PPB		14 Jun 2021 11:39:20	U	SMPL
M2672-01MS - 1	17-B6(8-12)MS	25194	4.8613	-7471B	PPB		14 Jun 2021 11:41:37	U	SMPL
M2672-01MSD - 1	17-B6(8-12)MSD	23185	4.4783	-7471B	PPB		14 Jun 2021 11:43:53	U	SMPL
M2672-02 - 1	17-B6(12-16)	1075	0.2625	-7471B	PPB		14 Jun 2021 11:46:09	U	SMPL
CCV88 - 1	CCV88	25714	4.9605	-7471B	PPB		14 Jun 2021 11:48:25	U	SMPL
CCB88 - 1	CCB88	-110	0.0366	-7471B	PPB		14 Jun 2021 11:50:40	U	SMPL
M2672-03 - 1	17-B6(16-20)	818	0.2135	-7471B	PPB		14 Jun 2021 11:52:56	U	SMPL
M2672-04 - 1	17-B6(20-24)	317	0.1180	-7471B	PPB		14 Jun 2021 11:55:12	U	SMPL
M2672-05 - 1	17-B6(24-28)	222	0.0999	-7471B	PPB		14 Jun 2021 11:57:27	U	SMPL
M2672-06 - 1	17-B6(28-32)	327	0.1199	-7471B	PPB		14 Jun 2021 11:59:43	U	SMPL
M2672-07 - 1	17-B6(32-36)	323	0.1192	-7471B	PPB		14 Jun 2021 12:01:58	U	SMPL
M2672-08 - 1	17-B6(36-40)	253	0.1058	-7471B	PPB		14 Jun 2021 12:04:14	U	SMPL
M2672-09 - 1	17-B6(40-44)	222	0.0999	-7471B	PPB		14 Jun 2021 12:06:30	U	SMPL
M2672-10 - 1	17-B6(44-48)	224	0.1003	-7471B	PPB		14 Jun 2021 12:08:47	U	SMPL
M2672-11 - 1	17-B6(48-52)	223	0.1001	-7471B	PPB		14 Jun 2021 12:11:04	U	SMPL
M2672-12 - 1	17-B6(52-56)	172	0.0904	-7471B	PPB		14 Jun 2021 12:13:19	U	SMPL
CCV89 - 1	CCV89	25716	4.9608	-7471B	PPB		14 Jun 2021 12:15:35	U	SMPL
CCB89 - 1	CCB89	-90	0.0404	-7471B	PPB		14 Jun 2021 12:17:50	U	SMPL
M2672-13 - 1	17-B6(56-60)	220	0.0995	-7471B	PPB		14 Jun 2021 12:20:06	U	SMPL
M2672-14 - 1	17-B6(60-64)	141	0.0845	-7471B	PPB		14 Jun 2021 12:22:21	U	SMPL
M2672-15 - 1	17-B6(64-68)	223	0.1001	-7471B	PPB		14 Jun 2021 12:24:37	U	SMPL
M2672-16 - 1	17-B6(68-72)	111	0.0787	-7471B	PPB		14 Jun 2021 12:26:52	U	SMPL
M2672-17 - 1	17-B6(72-76)	686	0.1884	-7471B	PPB		14 Jun 2021 12:29:08	U	SMPL
M2672-18 - 1	17-B6(76-80)	154	0.0869	-7471B	PPB		14 Jun 2021 12:31:24	U	SMPL
M2672-19 - 1	17-B6(80-84)	88	0.0743	-7471B	PPB		14 Jun 2021 12:33:41	U	SMPL
M2672-20 - 1	17-B6(84-88)	101	0.0768	-7471B	PPB		14 Jun 2021 12:35:57	U	SMPL
M2672-01LX5 - 1		568	0.1659	-7471B	PPB		14 Jun 2021 12:38:14	U	SMPL
CCV90 - 1	CCV90	24009	4.6354	-7471B	PPB		14 Jun 2021 12:40:30	U	SMPL
CCB90 - 1	CCB90	26201	5.0533	-7471B	PPB		14 Jun 2021 12:42:45	U	SMPL
		-121	0.0345	-7471B	PPB		14 Jun 2021 12:45:02	U	SMPL