



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114944

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	06/14/2021	10:47
2	CAL2	1	Water	NA	NA	20.3	7.00	06/14/2021	10:48
3	CAL3	1	Water	NA	NA	20.3	10.02	06/14/2021	10:49
4	ICV	1	Water	NA	NA	20.4	6.99	06/14/2021	10:50
5	CCV1	1	Water	NA	NA	20.3	2.01	06/14/2021	10:52
6	M2706-01	1	Water	NA	NA	20.1	8.58	06/14/2021	11:00
7	M2706-01DUP	1	Water	NA	NA	20.2	8.59	06/14/2021	11:01
8	CCV2	1	Water	NA	NA	20.4	12.02	06/14/2021	12:05

WORKLIST(Hardcopy Internal Chain)

VB 116945

WorkList Name : PH M2706 WorkList ID : 149826 Department : Wet-Chemistry Date : 06-14-2021 10:41:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/11/2021	Water	M2706-01	pH	Cool 4 deg C	ROYF02	F51	CIEW_WC	06/11/2021	9040C

Date/Time 06.14.21 10:45
 Raw Sample Received by: JB (wet 12:00)
 Raw Sample Relinquished by: CP SM

Date/Time 06.15.21 12:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JB wet 12:00