


LB114953

7470A

INSTRUMENT 1D, CV1



Linear

A= 0.0000e+000

B= 2.6536e-004

C= -8.1530e-002

Rho= 0.9993165

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.0	0.000	0.002	0.002	314	0.000	314					-
0.2	0.200	0.222	0.022	1145	0.0 %	1145					11
2.5	2.500	2.316	-0.184	9035	0.0 %	9035					-7
5.0	5.000	5.183	0.183	19839	0.0 %	19839					4
7.5	7.500	7.636	0.136	29083	0.0 %	29083					2
10.0	10.000	9.841	-0.159	37392	0.0 %	37392					-2

LB114953 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Jun 2021 10:00:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	314	-	0.0000	7470A	PPB	15 Jun 2021 10:18:39	S	Std
0.2 - 1	50.2	1145	-	0.2000	7470A	PPB	15 Jun 2021 10:20:56	S	Std
2.5 - 1	52.5	9035	-	2.5000	7470A	PPB	15 Jun 2021 10:23:13	S	Std
5.0 - 1	55	19839	-	5.0000	7470A	PPB	15 Jun 2021 10:25:29	S	Std
7.5 - 1	57.5	29083	-	7.5000	7470A	PPB	15 Jun 2021 10:27:46	S	Std
10.0 - 1	510	37392	-	10.0000	7470A	PPB	15 Jun 2021 10:30:02	S	Std
ICV63 - 1	ICV63	15695	4.0833	-	7470A	PPB	15 Jun 2021 10:35:16	U	SMPL
ICB63 - 1	ICB63	213	-0.0250	-	7470A	PPB	15 Jun 2021 10:40:03	U	SMPL
CCV95 - 1	CCV95	18447	4.8136	-	7470A	PPB	15 Jun 2021 10:42:17	U	SMPL
CCB95 - 1	CCB95	185	-0.0324	-	7470A	PPB	15 Jun 2021 10:47:40	U	SMPL
CRA - 1	CRA	983	0.1793	-	7470A	PPB	15 Jun 2021 10:49:55	U	SMPL
HighStd - 1	HighStd	35241	9.2701	-	7470A	PPB	15 Jun 2021 10:52:09	U	SMPL
ChkStd - 1	ChkStd	25496	6.6841	-	7470A	PPB	15 Jun 2021 10:54:25	U	SMPL
PB137098BL - 1	PBW	69	-0.0632	-	7470A	PPB	15 Jun 2021 10:56:42	U	SMPL
PB137098BS - 1	LCSW	16152	4.2046	-	7470A	PPB	15 Jun 2021 10:58:58	U	SMPL
M2639-02 - 1	B67	129	-0.0473	-	7470A	PPB	15 Jun 2021 11:01:14	U	SMPL
M2672-40 - 1	17-B6(84-88)	242	-0.0173	-	7470A	PPB	15 Jun 2021 11:03:30	U	SMPL
M2687-11 - 1	17-B6(100-104)	309	0.0005	-	7470A	PPB	15 Jun 2021 11:05:47	U	SMPL
M2687-12 - 1	17-B6(104-108)	316	0.0023	-	7470A	PPB	15 Jun 2021 11:08:04	U	SMPL
M2687-13 - 1	17-B6(108-112)	276	-0.0083	-	7470A	PPB	15 Jun 2021 11:10:19	U	SMPL
CCV96 - 1	CCV96	18280	4.7693	-	7470A	PPB	15 Jun 2021 11:12:34	U	SMPL
CCB96 - 1	CCB96	-62	-0.0980	-	7470A	PPB	15 Jun 2021 11:14:49	U	SMPL
M2687-14 - 1	17-B6(112-116)	269	-0.0101	-	7470A	PPB	15 Jun 2021 11:17:05	U	SMPL
M2687-15 - 1	17-B6(116-120)	302	-0.0014	-	7470A	PPB	15 Jun 2021 11:19:20	U	SMPL
M2687-16 - 1	17-B6(120-124)	370	0.0167	-	7470A	PPB	15 Jun 2021 11:21:35	U	SMPL
M2687-17 - 1	17-B6(124-125)	293	-0.0038	-	7470A	PPB	15 Jun 2021 11:23:50	U	SMPL
M2687-18 - 1	DUP-01	277	-0.0080	-	7470A	PPB	15 Jun 2021 11:26:06	U	SMPL
M2687-19 - 1	DUP-02	285	-0.0059	-	7470A	PPB	15 Jun 2021 11:28:22	U	SMPL
M2687-20 - 1	DUP-03	274	-0.0088	-	7470A	PPB	15 Jun 2021 11:30:38	U	SMPL
M2687-20DUP - 1	DUP-03DUP	297	-0.0027	-	7470A	PPB	15 Jun 2021 11:32:55	U	SMPL
M2687-20MS - 1	DUP-03MS	15346	3.9907	-	7470A	PPB	15 Jun 2021 11:35:12	U	SMPL
M2687-20MSD - 1	DUP-03MSD	15739	4.0950	-	7470A	PPB	15 Jun 2021 11:40:01	U	SMPL
CCV97 - 1	CCV97	18691	4.8784	-	7470A	PPB	15 Jun 2021 11:42:17	U	SMPL
CCB97 - 1	CCB97	86	-0.0587	-	7470A	PPB	15 Jun 2021 11:47:15	U	SMPL
PB137099BL - 1	PBW	272	-0.0094	-	7470A	PPB	15 Jun 2021 11:49:30	U	SMPL
PB137099BS - 1	LCSW	15328	3.9859	-	7470A	PPB	15 Jun 2021 11:54:20	U	SMPL
M2706-01 - 1	CIEW_WC	166	-0.0375	-	7470A	PPB	15 Jun 2021 11:56:35	U	SMPL
M2706-01DUP - 1	CIEW_WCDUP	286	-0.0056	-	7470A	PPB	15 Jun 2021 11:58:50	U	SMPL
M2706-01MS - 1	CIEW_WCMS	13530	3.5088	-	7470A	PPB	15 Jun 2021 12:01:06	U	SMPL
M2706-01MSD - 1	CIEW_WCMSD	13711	3.5569	-	7470A	PPB	15 Jun 2021 12:03:22	U	SMPL
CCV98 - 1	CCV98	19136	4.9964	-	7470A	PPB	15 Jun 2021 12:05:38	U	SMPL
CCB98 - 1	CCB98	136	-0.0454	-	7470A	PPB	15 Jun 2021 12:10:33	U	SMPL
PB137098TB - 1	PB137052TB	256	-0.0136	-	7470A	PPB	15 Jun 2021 12:12:49	U	SMPL
PB137099TB - 1	PB137087TB	310	0.0007	-	7470A	PPB	15 Jun 2021 12:15:05	U	SMPL
M2687-20LX5 - 1		58	-0.0661	-	7470A	PPB	15 Jun 2021 12:17:20	U	SMPL
M2687-20A - 1		15172	3.9445	-	7470A	PPB	15 Jun 2021 12:19:36	U	SMPL
M2706-01LX5 - 1		-56	-0.0964	-	7470A	PPB	15 Jun 2021 12:21:51	U	SMPL
M2706-01A - 1		15100	3.9254	-	7470A	PPB	15 Jun 2021 12:24:07	U	SMPL
CCV99 - 1	CCV99	18123	4.7276	-	7470A	PPB	15 Jun 2021 12:26:22	U	SMPL
CCB99 - 1	CCB99	40	-0.0709	-	7470A	PPB	15 Jun 2021 12:28:37	U	SMPL