



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114955

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	06/15/2021	09:40
2	M2706-01		1	9	100.00	>212.0	>212.0	06/15/2021	10:05
3	M2706-01DUP		1	9	100.00	>212.0	>212.0	06/15/2021	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB1149SS

WorkList Name : fp61521 WorkList ID : 149851 Department : Wet-Chemistry Date : 06-15-2021 09:04:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/21/2021	Water	M2706-01	F	Cool 4 deg C	ROYF02	F51	CIEW_WC	06/11/2021	1010A

Date/Time 06/15/21 09:10
Raw Sample Received by: KS (Weber)
Raw Sample Relinquished by: JD CSM

Date/Time 06/15/21 10:40
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: KS (Weber)