



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114967

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/15/2021	16:50
2	CAL2	1	Water	NA	NA	20.7	7.01	06/15/2021	16:51
3	CAL3	1	Water	NA	NA	20.6	10.03	06/15/2021	16:52
4	ICV	1	Water	NA	NA	20.6	6.99	06/15/2021	17:00
5	CCV1	1	Water	NA	NA	20.3	2.01	06/15/2021	17:02
6	M2714-03	1	Water	NA	NA	23.7	5.58	06/15/2021	17:05
7	M2714-05	1	Water	NA	NA	23.8	5.41	06/15/2021	17:06
8	M2714-07	1	Water	NA	NA	23.9	5.14	06/15/2021	17:10
9	M2716-03	1	Water	NA	NA	23.1	6.79	06/15/2021	17:11
10	M2716-05	1	Water	NA	NA	23.3	8.06	06/15/2021	17:12
11	M2717-01	1	Water	NA	NA	23.6	5.99	06/15/2021	17:15
12	M2717-01DUP	1	Water	NA	NA	23.7	6.00	06/15/2021	17:16
13	CCV2	1	Water	NA	NA	20.6	12.02	06/15/2021	17:19

VB 114967

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2714 W WorkList ID : 149862 Department : Wet-Chemistry Date : 06-15-2021 16:44:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/15/2021	Water	M2714-03	pH	Cool 4 deg C	PSEG03	G42	33225	06/15/2021	9040C
06/15/2021	Water	M2714-05	pH	Cool 4 deg C	PSEG03	G42	33236	06/15/2021	9040C
06/15/2021	Water	M2714-07	pH	Cool 4 deg C	PSEG03	G42	33242	06/15/2021	9040C
06/15/2021	Water	M2716-03	pH	Cool 4 deg C	PSEG03	G42	33219	06/15/2021	9040C
06/15/2021	Water	M2716-05	pH	Cool 4 deg C	PSEG03	G42	33220	06/15/2021	9040C
06/15/2021	Water	M2717-01	pH	Cool 4 deg C	PSEG03	G41	33233	06/15/2021	9040C

Date/Time 06-15-21 16:45
Raw Sample Received by: JP Wet Chem
Raw Sample Relinquished by: R-C (sam)

Date/Time 06-15-21 18:10
Raw Sample Received by: R-C (sam)
Raw Sample Relinquished by: JP Wet Chem