



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114968

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/15/2021	16:50
2	CAL2	1	Water	NA	NA	20.7	7.01	06/15/2021	16:51
3	CAL3	1	Water	NA	NA	20.6	10.03	06/15/2021	16:52
4	ICV	1	Water	NA	NA	20.6	6.99	06/15/2021	17:00
5	CCV1	1	Water	NA	NA	20.3	2.01	06/15/2021	17:02
6	M2708-01	1	Solid	20.03	20	22.4	6.84	06/15/2021	17:20
7	M2714-01	1	Solid	20.02	20	23.7	5.58	06/15/2021	17:21
8	M2715-01	1	Solid	20.03	20	23.4	5.63	06/15/2021	17:22
9	M2716-01	1	Solid	20.04	20	23.3	5.98	06/15/2021	17:24
10	M2716-01DUP	1	Solid	20.03	20	23.4	5.99	06/15/2021	17:25
11	CCV2	1	Water	NA	NA	20.3	12.02	06/15/2021	17:26

LB 114968

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2716 S

WorkList ID : 149863

Department : Wet-Chemistry

Date : 06-15-2021 16:45:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/14/2021	Solid	M2708-01	pH	Cool 4 deg C	PSEG05	G41	AU-06-061421	06/14/2021	9045D
06/15/2021	Solid	M2714-01	pH	Cool 4 deg C	PSEG03	G42	33241	06/15/2021	9045D
06/15/2021	Solid	M2715-01	pH	Cool 4 deg C	PSEG03	G42	33248	06/15/2021	9045D
06/15/2021	Solid	M2716-01	pH	Cool 4 deg C	PSEG03	G42	33254	06/15/2021	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: