



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114974

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	06/16/2021	09:35
2	M2715-03		1	3	22.00	71.6	71.9	06/16/2021	10:05
3	M2715-03DUP		1	3	21.00	69.8	70.1	06/16/2021	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114974

WorkList Name : fp61621-1 WorkList ID : 149875 Department : Wet-Chemistry Date : 06-16-2021 09:05:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/22/2021	Solid	M2715-03	Flash Point	Cool 4 deg C	PSEG03	G42	33258	06/15/2021	1010A

Date/Time 06/16/21 09:10
 Raw Sample Received by: KS (Weber)
 Raw Sample Relinquished by: CL SM

Date/Time 06/16/21 10:45
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: KS (Weber)