



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114976

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	06/16/2021	12:25
2	M2714-03		1	9	100.00	>212.0	>212.0	06/16/2021	12:45
3	M2714-03DUP		1	9	100.00	>212.0	>212.0	06/16/2021	13:10
4	M2714-05		1	8	100.00	>212.0	>212.0	06/16/2021	13:35
5	M2714-07		1	9	100.00	>212.0	>212.0	06/16/2021	14:05
6	M2716-03		1	9	100.00	>212.0	>212.0	06/16/2021	14:40
7	M2716-05		1	8	100.00	>212.0	>212.0	06/16/2021	15:05
8	M2717-01		1	8	100.00	>212.0	>212.0	06/16/2021	15:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 114976

WorkList Name : fp61621

WorkList ID : 149874

Department : Wet-Chemistry

Date : 06-16-2021 09:04:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/22/2021	Water	M2714-03 <u>B</u>	Flash Point	Cool 4 deg C	PSEG03	G42	33225	06/15/2021	1010A
06/22/2021	Water	M2714-05 <u>C</u>	Flash Point	Cool 4 deg C	PSEG03	G42	33236	06/15/2021	1010A
06/22/2021	Water	M2714-07 <u>B</u>	Flash Point	Cool 4 deg C	PSEG03	G42	33242	06/15/2021	1010A
06/22/2021	Water	M2716-03 <u>C</u>	Flash Point	Cool 4 deg C	PSEG03	G42	33219	06/15/2021	1010A
06/22/2021	Water	M2716-05 <u>B</u>	Flash Point	Cool 4 deg C	PSEG03	G42	33220	06/15/2021	1010A
06/22/2021	Water	M2717-01 <u>D</u>	Flash Point	Cool 4 deg C	PSEG03	G41	33233	06/15/2021	1010A

Date/Time 06/16/21 12:00
 Raw Sample Received by: RS (Wetchem)
 Raw Sample Relinquished by: CP SM

Date/Time 06/16/21 15:40
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: RS (Wetchem)