

LB114978


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37682



A= 0.0000e+000

B= 2.6578e-004 slope

C= 6.2060e-004 y-intercept

Rho= 0.9999764

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.002	0.002	7	0.000	7					1
0.05	0.050	0.046	-0.004	170	0.0 %	170					8
0.20	0.200	0.207	0.007	778	0.0 %	778					3
2.50	2.500	2.479	-0.021	9324	0.0 %	9324					1
5.00	5.000	5.043	0.043	18974	0.0 %	18974					1
7.50	7.500	7.456	-0.044	28053	0.0 %	28053					1
10.0	10.000	10.016	0.016	37682	0.0 %	37682					0

LB114978
INSTRUMENT ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		7 -		0 SFAM01.1	PPB	6/16/2021 12:22	Std
0.05 S01		170 -		0.05 SFAM01.1	PPB	6/16/2021 12:24	Std
0.2 S02		778 -		0.2 SFAM01.1	PPB	6/16/2021 12:27	Std
2.5 S03		9324 -		2.5 SFAM01.1	PPB	6/16/2021 12:29	Std
5 S04		18974 -		5 SFAM01.1	PPB	6/16/2021 12:31	Std
7.5 S05		28053 -		7.5 SFAM01.1	PPB	6/16/2021 12:34	Std
10 S06		37682 -		10 SFAM01.1	PPB	6/16/2021 12:36	Std
ICV043	ICV043	14805	3.9355 -	SFAM01.1	PPB	6/16/2021 12:39	SMPL
ICB043	ICB043	-152	-0.0398 -	SFAM01.1	PPB	6/16/2021 12:41	SMPL
CCV013	CCV013	18741	4.9816 -	SFAM01.1	PPB	6/16/2021 12:43	SMPL
CCB013	CCB013	-138	-0.0361 -	SFAM01.1	PPB	6/16/2021 12:45	SMPL
PB137072BL	PBW072	-10	-0.002 -	SFAM01.1	PPB	6/16/2021 12:48	SMPL
M2250-02	MDL-WATE	167	0.045 -	SFAM01.1	PPB	6/16/2021 12:52	SMPL
BLANK-06	BLANK-06	6	0.0022 -	SFAM01.1	PPB	6/16/2021 12:55	SMPL
CCV014	CCV014	18997	5.0496 -	SFAM01.1	PPB	6/16/2021 12:57	SMPL
CCB014	CCB014	-133	-0.0347 -	SFAM01.1	PPB	6/16/2021 12:59	SMPL