



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114979

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	06/16/2021	13:50
2	CAL2	1	Water	NA	NA	20.6	7.00	06/16/2021	13:51
3	CAL3	1	Water	NA	NA	10.03	10.02	06/16/2021	13:52
4	ICV	1	Water	NA	NA	20.6	6.99	06/16/2021	14:00
5	CCV1	1	Water	NA	NA	20.6	2.01	06/16/2021	14:01
6	M2743-01	1	Water	NA	NA	23.9	5.05	06/16/2021	14:03
7	M2743-01DUP	1	Water	NA	NA	23.8	5.06	06/16/2021	14:04
8	CCV2	1	Water	NA	NA	20.6	12.02	06/16/2021	14:05

2B 114979

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2743 WorkList ID : 149881 Department : Wet-Chemistry Date : 06-16-2021 13:40:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/11/2021	Water	M2743-01	pH	Cool 4 deg C	STER17	F62	JS9600	06/11/2021	9040C

Date/Time 06.16.21 13:45
Raw Sample Received by: JP
Raw Sample Relinquished by: CP SM

Date/Time 06-16-21 14:45
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JP Out (20m)