

LB114983


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41279



A= 0.0000e+000

B= 2.4041e-004 slope

C= -5.6128e-003 y-intercept

Rho= 0.9999061

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.007	-0.007	-4	0.000	-4					5
0.05	0.050	0.040	-0.010	191	0.0 %	191					20
0.20	0.200	0.200	0.000	857	0.0 %	857					0
2.50	2.500	2.463	-0.037	10270	0.0 %	10270					7
5.00	5.000	5.046	0.046	21013	0.0 %	21013					1
7.50	7.500	7.588	0.088	31588	0.0 %	31588					1
10.0	10.000	9.918	-0.082	41279	0.0 %	41279					1

LB114983
INSTRUMENT ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		-4 -		0 SFAM01.1	PPB	6/16/2021 14:28	Std
0.05 S01	191	-	0.05	SFAM01.1	PPB	6/16/2021 14:30	Std
0.2 S02	857	-	0.2	SFAM01.1	PPB	6/16/2021 14:32	Std
2.5 S03	10270	-	2.5	SFAM01.1	PPB	6/16/2021 14:34	Std
5 S04	21013	-	5	SFAM01.1	PPB	6/16/2021 14:37	Std
7.5 S05	31588	-	7.5	SFAM01.1	PPB	6/16/2021 14:39	Std
10 S06	41279	-	10	SFAM01.1	PPB	6/16/2021 14:41	Std
ICV044	ICV044	16645	3.9959 -	SFAM01.1	PPB	6/16/2021 14:44	SMPL
ICB044	ICB044	-27	-0.0121 -	SFAM01.1	PPB	6/16/2021 14:46	SMPL
CCV015	CCV015	21025	5.0489 -	SFAM01.1	PPB	6/16/2021 14:48	SMPL
CCB015	CCB015	-21	-0.0107 -	SFAM01.1	PPB	6/16/2021 14:51	SMPL
PB137152E	PBS152	-8	-0.0075 -	SFAM01.1	PPB	6/16/2021 14:53	SMPL
M2250-04	MDL-SOIL-(	201	0.0427 -	SFAM01.1	PPB	6/16/2021 14:58	SMPL
BLANK-08	BLANK-08	3	-0.0049 -	SFAM01.1	PPB	6/16/2021 15:00	SMPL
CCV016	CCV016	21036	5.0516 -	SFAM01.1	PPB	6/16/2021 15:02	SMPL
CCB016	CCB016	-52	-0.0181 -	SFAM01.1	PPB	6/16/2021 15:05	SMPL