


LB114986

SFAM01.1

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.6408e-004 *slope*C= -8.9014e-003 *y-intercept*

Rho= 0.9999809

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.001	0.001	39	0.000	39					-
0.05	0.050	0.046	-0.004	207	0.0 %	207					8
0.20	0.200	0.206	0.006	815	0.0 %	815					3
2.50	2.500	2.529	0.029	9611	0.0 %	9611					1
5.00	5.000	4.961	-0.039	18820	0.0 %	18820					1
7.50	7.500	7.478	-0.022	28352	0.0 %	28352					0
10.0	10.000	10.028	0.028	38007	0.0 %	38007					0

LB114986

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	39	-		0 SFAM01.1	PPB	6/16/2021 15:53	Std
	0.05 S01	207	-	0.05	SFAM01.1	PPB	6/16/2021 15:56	Std
	0.2 S02	815	-	0.2	SFAM01.1	PPB	6/16/2021 15:58	Std
	2.5 S03	9611	-	2.5	SFAM01.1	PPB	6/16/2021 16:00	Std
	5 S04	18820	-	5	SFAM01.1	PPB	6/16/2021 16:02	Std
	7.5 S05	28352	-	7.5	SFAM01.1	PPB	6/16/2021 16:05	Std
	10 S06	38007	-	10	SFAM01.1	PPB	6/16/2021 16:07	Std
ICV045	ICV045	14973	3.9452 -		SFAM01.1	PPB	6/16/2021 16:10	SMPL
ICB045	ICB045	33	-0.0002 -		SFAM01.1	PPB	6/16/2021 16:15	SMPL
CCV017	CCV017	18654	4.9172 -		SFAM01.1	PPB	6/16/2021 16:18	SMPL
CCB017	CCB017	23	-0.0028 -		SFAM01.1	PPB	6/16/2021 16:22	SMPL
PB137149BL	PBW149	22	-0.0031 -		SFAM01.1	PPB	6/16/2021 16:25	SMPL
M2606-01	EW116	62	0.0075 -		SFAM01.1	PPB	6/16/2021 16:27	SMPL
M2606-02	EW118	36	0.0006 -		SFAM01.1	PPB	6/16/2021 16:29	SMPL
M2606-03	EW118D	37	0.0009 -		SFAM01.1	PPB	6/16/2021 16:31	SMPL
M2606-04	EW118S	3613	0.9452 -		SFAM01.1	PPB	6/16/2021 16:34	SMPL
CCV018	CCV018	17562	4.6289 -		SFAM01.1	PPB	6/16/2021 16:36	SMPL
CCB018	CCB018	24	-0.0026 -		SFAM01.1	PPB	6/16/2021 16:41	SMPL