



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114990

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	06/16/2021	16:00
2	CAL2	1	Water	NA	NA	20.6	7.00	06/16/2021	16:01
3	CAL3	1	Water	NA	NA	20.6	10.03	06/16/2021	16:02
4	ICV	1	Water	NA	NA	20.6	7.00	06/16/2021	16:05
5	CCV1	1	Water	NA	NA	20.3	2.01	06/16/2021	16:06
6	M2710-02	1	Solid	20.02	20	22.7	7.24	06/16/2021	16:10
7	M2746-03	1	Solid	20.03	20	21.9	8.50	06/16/2021	16:11
8	M2746-03DUP	1	Solid	20.04	20	22.0	8.51	06/16/2021	16:12
9	CCV2	1	Water	NA	NA	20.6	12.02	06/16/2021	16:15

WORKLIST(Hardcopy Internal Chain)

28 114990

WorkList Name : CORROSIONITY M2715

WorkList ID : 149882

Department : Wet-Chemistry

Date : 06-16-2021 15:09:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/14/2021	Solid	M2710-02	Corrosivity	Cool 4 deg C	PENN11	F62	WC-1	06/14/2021	9045D
06/16/2021	Solid	M2746-03	Corrosivity	Cool 4 deg C	JPCL01	E62	NEW-DORP-2	06/16/2021	9045D

Date/Time 06.16.21 15:10
 Raw Sample Received by: JD (not (M2715))
 Raw Sample Relinquished by: JD (not (M2715))

Date/Time 06.16.21 17:20
 Raw Sample Received by: JD out (M2715)
 Raw Sample Relinquished by: JD out (M2715)