



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115000

Slope : 97.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/17/2021	11:05
2	CAL2	1	Water	NA	NA	20.3	7.00	06/17/2021	11:06
3	CAL3	1	Water	NA	NA	20.3	10.03	06/17/2021	11:07
4	ICV	1	Water	NA	NA	20.3	6.99	06/17/2021	11:08
5	CCV1	1	Water	NA	NA	20.3	2.01	06/17/2021	11:10
6	M2747-01	1	Solid	20.03	20	21.0	8.93	06/17/2021	11:13
7	M2747-01DUP	1	Solid	20.04	20	21.1	8.94	06/17/2021	11:14
8	M2747-02	1	Solid	20.02	20	20.9	8.52	06/17/2021	11:15
9	M2747-03	1	Solid	20.02	20	20.8	8.26	06/17/2021	11:16
10	M2747-04	1	Solid	20.03	20	20.6	8.42	06/17/2021	11:19
11	M2747-05	1	Solid	20.02	20	21.1	8.58	06/17/2021	11:20
12	M2747-06	1	Solid	20.03	20	20.9	8.62	06/17/2021	11:22
13	M2747-07	1	Solid	20.05	20	20.8	8.39	06/17/2021	11:25
14	M2747-08	1	Solid	20.02	20	20.5	8.51	06/17/2021	11:26
15	M2747-09	1	Solid	20.03	20	20.6	8.72	06/17/2021	11:28
16	CCV2	1	Water	NA	NA	20.4	12.02	06/17/2021	11:30
17	M2747-10	1	Solid	20.02	20	20.5	8.52	06/17/2021	11:32
18	M2747-11	1	Solid	20.01	20	20.0	8.79	06/17/2021	11:34
19	M2747-12	1	Solid	20.03	20	20.0	8.67	06/17/2021	11:35
20	CCV3	1	Water	NA	NA	20.4	2.01	06/17/2021	11:36

115000

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2747 WorkList ID : 149918 Department : Wet-Chemistry Date : 06-17-2021 10:54:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/20/2021	Solid	M2747-01	pH	Cool 4 deg C	GEIC06	F62	A1(WINDROW-LEFT-OF-	05/20/2021	9045D
05/20/2021	Solid	M2747-02	pH	Cool 4 deg C	GEIC06	F62	A2(WINDROW-MIDDLE)	05/20/2021	9045D
05/20/2021	Solid	M2747-03	pH	Cool 4 deg C	GEIC06	F62	A3(WINDROW-RIGHT)	05/20/2021	9045D
05/20/2021	Solid	M2747-04	pH	Cool 4 deg C	GEIC06	F62	X4(TOPSOIL-PILE)	05/20/2021	9045D
05/27/2021	Solid	M2747-05	pH	Cool 4 deg C	GEIC06	F62	B1(WINDROW-1-ABOVE	05/27/2021	9045D
05/27/2021	Solid	M2747-06	pH	Cool 4 deg C	GEIC06	F62	B2(WINDROW-2-ABOVE	05/27/2021	9045D
05/27/2021	Solid	M2747-07	pH	Cool 4 deg C	GEIC06	F62	B3(WINDROW-3-ABOVE	05/27/2021	9045D
05/27/2021	Solid	M2747-08	pH	Cool 4 deg C	GEIC06	F62	X5(TOPSOIL-PILE)	05/27/2021	9045D
06/10/2021	Solid	M2747-09	pH	Cool 4 deg C	GEIC06	F62	C1(WINDROW-LEFT-OF-	06/10/2021	9045D
06/10/2021	Solid	M2747-10	pH	Cool 4 deg C	GEIC06	F62	C2(WINDROW-MIDDLE)	06/10/2021	9045D
06/10/2021	Solid	M2747-11	pH	Cool 4 deg C	GEIC06	F62	C33(WINDROW-RIGHT)	06/10/2021	9045D
06/10/2021	Solid	M2747-12	pH	Cool 4 deg C	GEIC06	F62	X6(TOPSOIL-PILE)	06/10/2021	9045D

Date/Time 06.17.21 11:00
 Raw Sample Received by: JB (wet-chem)
 Raw Sample Relinquished by: AM (wet-chem)

Date/Time 06.17.21 12:35
 Raw Sample Received by: AM (wet-chem)
 Raw Sample Relinquished by: JB (wet-chem)