



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115002

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/17/2021	12:05
2	CAL2	1	Water	NA	NA	20.3	7.00	06/17/2021	12:06
3	CAL3	1	Water	NA	NA	20.3	10.03	06/17/2021	12:07
4	ICV	1	Water	NA	NA	20.3	7.00	06/17/2021	12:08
5	CCV1	1	Water	NA	NA	20.3	2.01	06/17/2021	12:10
6	M2762-02	1	Water	NA	NA	20.8	9.79	06/17/2021	12:20
7	M2762-02DUP	1	Water	NA	NA	20.9	9.80	06/17/2021	12:21
8	CCV2	1	Water	NA	NA	20.4	12.02	06/17/2021	12:22

2B 115002

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S M2710 WorkList ID : 149923 Department : Wet-Chemistry Date : 06-17-2021 11:53:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/17/2021	Water	M2762-02	pH	Cool 4 deg C	ROYF02	F52	WC_AQUEOUS_DBD	06/17/2021	9040C

Date/Time 06-17-21 12:00

Raw Sample Received by: JP [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06-17-21 13:00

Raw Sample Received by: J.C. Sun

Raw Sample Relinquished by: JP [Signature]